



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 001

#### PROCUREMENT OF GLUCOMETER STRIPS (SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Packaging – Manufacturer Standard With glucometer compatible to strips with battery	
<b>FROM</b>	<b>TO</b>
Assay Range: 0.50-33.30 mmol/L (9-600 mg/dL)	Assay Range: 0.50 - 1.1 up to 33.30 mmol/L (9-19.8 up to 600 mg/dL)
Schedule of delivery: 30-90 days	Schedule of delivery: 30-60 days


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 28<sup>th</sup> day of November 2023, 9:30 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 18<sup>th</sup> day of November 2022 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson *4*



Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>PROCUREMENT OF GLUCOMETER STRIPS</b>	Qty./Unit	11,110 CANISTERS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC: P 3,333,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b><u>SPECIFICATIONS:</u></b></p> <p>Wide Gold Electrode Strips – 25 strips per box/canister            With one glucometer for every 500 strips            With one glucometer control solution for every 500 strips            Assay Range: 0.50 - 1.1 up to 33.30 mmol/L (9-19.8 up to 600 mg/dL)            Test Blood Volume: 0.5-1 uL            Sample Types: Capillary/Venous whole blood            With glucometer compatible to strips with battery            Shelf Life: at least 18 months upon delivery            Packaging – Manufacturer Standard</p> <p>On each canister/box, the following should be legibly imprinted or stickered with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:</p> <p>Philippine Government Property- Department of Health            NOT FOR SALE</p> <p>Area of delivery: DOH-MMCHD Pasig Warehouse            Schedule of delivery: 30-60 days</p>			
<p><b><u>ADDITIONAL REQUIREMENT TO BE SUBMITTED BY THE SINGLE/LOWEST CALCULATED BIDDER</u></b></p>			



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

**(SCB/LCB) AS PART OF POST QUALIFICATION:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.