



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 018

PROCUREMENT OF DENGUE NS1 AG + IGG/IGM RAPID DIAGNOSTIC TEST (SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<u>SPECIFICATIONS:</u>	
DENGUE NS1 AG + IgG/IgM RAPID DIAGNOSTIC TEST	
Delivery site: DOH-MMCHD Tala Warehouse	
FROM	TO

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 28th day of November 2022, 9:30 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 18th day of November 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson *y*



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	PROCUREMENT OF DENGUE NS1 AG + IGG/IGM RAPID DIAGNOSTIC TEST (SHORT OF AWARD)	Qty./Unit	1,462 KITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: P 5,848,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>SPECIFICATIONS:</u></p> <p>DENGUE NS1 AG + IgG/IgM RAPID DIAGNOSTIC TEST</p> <p>Standard Features:</p> <ul style="list-style-type: none"> *RDT kit for the qualitative detection of dengue non-structural protein 1 (NS1) and dengue Ns1 IgG and IgM antibodies from all four (4) Dengue serotypes *NS1 & IgG/IgM Combo in Immunochromatography in plastic cassette format *3mL-5mL assay buffer *10uL capillary pipette *1 disposable dropper *1 lancet, ultra thin, 28 gauge, consistent depth penetration, universal design fits almost all lancing devices *assay time 15-20 min *Sensitivity: $\geq 92\%$ and Specificity $\geq 98\%$ Dengue NS1 Ag based on 3rd party evaluation or based on publication by National Reference Laboratory (NRL), Centers for Disease Control and Prevention (CDC) or World Health Organization (WHO) *Sensitivity $\geq 94\%$ and Specificity Dengue IgG/gM $\geq 96\%$ Dengue IgG/IgM based on 3rd party evaluation or based on publication by National Reference Laboratory (NRL), Centers for Disease Control and Prevention (CDC) or World Health Organization (WHO). 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>*Packaging: Manufacturer's standard</p> <p>Recall and Replacement:</p> <p>1. The Supplier must ensure the quality of products and if there will be problems in the quality, the Supplier will recall and replace the products distributed in the hospitals, health centers and public health laboratories</p> <p>*Delivery Time: 60 days after receipt of PO</p> <p>*Shelf life: Products must have a minimum shelf life of eighteen (18) months remaining at the time of delivery.</p> <p>Delivery site: DOH-MMCHD Tala Warehouse</p>	
<p><u>ADDITIONAL REQUIREMENT TO BE SUBMITTED BY THE SINGLE/LOWEST CALCULATED BIDDER (SCB/LCB) AS PART OF POST QUALIFICATION:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	