

Republic of the Philippines Department of Health METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 022E

PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:		
Add:		
>Existing tower: brand: OUMAN		
>Model: 910HDA		
FROM	ТО	
Delivery:	Delivery:	
> 30 calendar days.	> 30-60 calendar days.	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25th day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of May 2023 in MMCHD

Approved by:

SGD PRETCHELL P. TOLENTINO, MD, MCHM, Director III / BAC Chairperson RUCOTHERHILAR

Republic of the Philippines	Republic of the Philippines			
Department of Health				
Metro Manila Center for Health Development				
TECHNICAL SPECIFICATIONS				
Item No. 1 LAPAROSCOPIC MACHINE INSTRUMENTS WITH MEDICAL RECORDER	Qty./Unit	1 unit		
Name of Manufacturer:	Country of Origin			
Brand:	Model: (if app	olicable)		
ABC: 1,500,000.00				
PURCHASER'S SPECIFICATION	STATEMENT COMPLIANC			
 A. SPECIFICATIONS: The following instruments must be compatible with the existing tower: Existing tower: Brand: OUMAN Model: 910HDA Laparoscopic Instruments 1 pc Laparoscopy Scope 10mm dia x 30 degree. 330mm 2 pcs Trocar sleeve with stop-cock & obturator pyramidal 6mm 1 pc Trocar sleeve with stop-cock & obturator Pyramidal 11mm 2 pcs Hasson Trocar complete 11mm 1 pc Reducer removable with holder, from 10mm - 5mm 1 pc Veress needle 2mm length 200mm 1 pc Karyland dissecting forceps 5mm 1 pc Blunt dissecting and grasping forceps 5mm 1 pc Clip Applying forceps 10mm 1 pc Bobcock grasping forceps 3aw 5mm 1 pc Metzenbaum scisor slightly curved 1 pc Endo needle holder 5mm 1 pc Lhook 5mm 1 pc L hook 5mm 1 pc Look 5mm 1 pc Spatula 5mm 1 pc Suction irrigator with sliding valve and standard tube 1 box (20 per box) Titanium clips 20 cannister instrument wipes containing: Belzankonium chloride - 0.310% 				

Republic of the Philippines Department of Health METRO MANILA CENTER FOR HEALTH DEVELOPMENT



- 1 minute effectivity upon contact	
Medical Recorder	
• MPEG-4AVC/H.264 G-4 AVC/H.264	
 HIGH QUALITY FULL HD record (MPEG-4 AVC/H.264 zip) 	
• Long recording time	
 Multiport (HDMI/DV, CVBS, YPbPr) 	
 Simplified keys and easy operation 	
Still and moving picture capture	
Interface connection	
-Interface: USB2.0	
-Video input: Digital- HD-SDI and Analog- CVBS (PAL)	
-Video output HDMI/DVI/VGA	
Hardware Specification Encoder II 264 Hardware encoding	
-Encoder: H.264 Hardware encoding	
-User interface: Recorder button. Capture button. Video source switching button. (Light indicator)	
-Storage Media: U Disk, 2T Max Capacity Mobile HDD	
-Storage Metha. O Disk, 21 Max Capacity Mobile IDD	
B. REQUIREMENTS IF AWARDED THE CONTRACT	
1. Completion Period: The delivery, installation, testing and commissioning of the	
equipment and its accessories, including the training of end-users and	
maintenance staff must be completed with <i>30-60 calendar days</i> upon receipt of	
Notice to Proceed.	
2. Testing: Prior to acceptance, the end user shall conduct a physical inspection	
and functionality test. The equipment must be functioning and must have no physical damage and defect.	
physical damage and delect.	
3. Training: The supplier shall provide a training on the proper use and	
maintenance of the equipment to the end-users and to the hospital maintenance	
staff within 3 days upon the delivery of the equipment.	
4. Warranty: Warranty certificate for two (2) years on parts and service. The	
supplier shall either repair or replace any item or part in the equipment that is	
found to be defective in material or in workmanship under normal use. The	
warranty period shall commence from the date of acceptance by the end-user	
after testing and commissioning.	
5. Notarized undertaking that the supplier shall conduct the necessary corrective	
maintenance within five (5) calendar days upon notification of the equipment	
breakdown from the end-user. The undertaking shall include a statement that the	
number of days where the equipment is unusable due to defective material or	
workmanship, shall be added to the warranty period.	
6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each	
unit.	
x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-	
Source of Fund: SAA 2022-02-0531 (HFEP 2022 ConAp)	
Recipients: Ospital ng Paranaque I	