



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 025E

PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Delivery Schedule: > 30 calendar days (CD)	Delivery Schedule: > 30-60 calendar days (CD)

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25th day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of May 2023 in MMCHD

Approved by:

SGD

PRETCHELL P. TOLENTINO, MD, MCHM,
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Republic of the Philippines</p> <p>Department of Health</p> <p>Metro Manila Center for Health Development</p> <p>TECHNICAL SPECIFICATIONS</p>															
Item No. 1	PULSE OXIMETER WITH ADULT AND PEDIA PROBE	Qty./Unit	14Units												
Name of Manufacturer:		Country of Origin													
Brand:		Model: (if applicable)													
ABC: P 980,000.00															
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE													
<p><u>A. SPECIFICATIONS:</u></p> <ul style="list-style-type: none"> • Type: Handheld (Portable) • Blood pressure rate, Oxygen saturation, Pulse rate, Temperature • At least 3.5" colored, high brightness LCD • LCD display brightness is adjustable • Resolution: at least 320*480 • Anti-motion artifacts, anti-perfusion • Large Font and Waveform display • Dynamic data and Graphic trends to view • Applies to adults, children and newborn • Trend Storage: at least 72 hours • Audio and Visual Alarm • Alarm: Probe off, Finger out, Low power • NIBP (Non-invas Blood Pressure) Measuring range: <table style="width: 100%; border: none;"> <tr> <td style="width: 33%;">Adult:</td> <td style="width: 33%;">PED:</td> <td style="width: 33%;">NEO:</td> </tr> <tr> <td>SYS: 40 ~ 270mmHg</td> <td>40 ~ 200mmHg</td> <td>40 ~ 135mmHg</td> </tr> <tr> <td>MEAN: 20 ~ 230mmHg</td> <td>20 ~ 165mmHg</td> <td>20 ~ 105mmHg</td> </tr> <tr> <td>DIA: 10 ~ 210mmHg</td> <td>10 ~ 150mmHg</td> <td>10 ~ 95mmHg</td> </tr> </table>		Adult:	PED:	NEO:	SYS: 40 ~ 270mmHg	40 ~ 200mmHg	40 ~ 135mmHg	MEAN: 20 ~ 230mmHg	20 ~ 165mmHg	20 ~ 105mmHg	DIA: 10 ~ 210mmHg	10 ~ 150mmHg	10 ~ 95mmHg		
Adult:	PED:	NEO:													
SYS: 40 ~ 270mmHg	40 ~ 200mmHg	40 ~ 135mmHg													
MEAN: 20 ~ 230mmHg	20 ~ 165mmHg	20 ~ 105mmHg													
DIA: 10 ~ 210mmHg	10 ~ 150mmHg	10 ~ 95mmHg													



<p>PULSE RATE: 40~240bpm 40~240bpm 40 ~ 240bpm</p> <ul style="list-style-type: none"> -Blood pressure precision average deviation: $\leq \pm 5\text{mmHg}$ -Blood pressure precision standard deviation: $\leq 8\text{mmHg}$ • SPO2 Pulse Oxygen Saturation: <ul style="list-style-type: none"> -Resolution: 1% -Accuracy: $\pm 2\text{bpm}$ (non-motion) and $\pm 3\text{bpm}$ (motion) -Pulse rate: 30-250 bpm • Temperature: <ul style="list-style-type: none"> - Range: 25 - 45°C - Resolution: 0.1°C - Sensitivity: $\pm 0.1^\circ\text{C}$ • Power: <ul style="list-style-type: none"> - Lithium Battery, operation time at least 15 hours for normal operation - Charger: built as a charging stand - 220V, 50/60Hz • Inclusion: <ul style="list-style-type: none"> - adult and pediatric reusable probe - adult cuff 	
--	--

<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <i>30-60 calendar days</i> upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period. 6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following: <ol style="list-style-type: none"> a) Service manual in English language 	
--	--

