



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2023 – 032E**

### PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
<b>Delivery:</b> >30 calendar days.	<b>Delivery:</b> >30-60 calendar days.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19<sup>th</sup> day of May 2023 in MMCHD

Approved by:

SGD

**PRETCHELL P. TOLENTINO, MD, MCHM,**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p><b>Republic of the Philippines</b></p> <p><b>Department of Health</b></p> <p><b>Metro Manila Center for Health Development</b></p> <p><b>TECHNICAL SPECIFICATIONS</b></p>			
Item No. 1	<b>AUTOMATED EXTERNAL DEFIBRILLATOR (AED)</b>	Qty./Unit	5 units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC: 750,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p><b><u>A. SPECIFICATIONS:</u></b></p> <ul style="list-style-type: none"> <li>• Two-button Operation: On/Off, Discharge (shock button)</li> <li>• Voice Prompts: Voice Command that gives instruction to the user in operating the device</li> <li>• Automatically evaluates patient impedance for proper pad contact.</li> <li>• Automatically analyses patient condition for shockable or non-shockable rhythms</li> <li>• Patient ECG can be acquired through defibrillator pads</li> <li>• Energy output accuracy: <math>\pm 15\%</math></li> <li>• Pulse shape: Biphasic</li> <li>• Charge time: maximum of 15 seconds</li> <li>• Battery Capacity: Capable of providing at least 200 discharges at maximum energy</li> <li>• Replaceable battery</li> <li>• Visible and Audible Indicators:               <ul style="list-style-type: none"> <li>- Low battery</li> <li>- Pad skin contact/disconnection</li> <li>- AED Status</li> </ul> </li> <li>• Defibrillator Pads: self-adhesive, disposable and non-polarized               <ul style="list-style-type: none"> <li>- at least five (5) pairs for adult</li> <li>- at least five (5) pairs for pedia</li> </ul> </li> </ul>			
<p><b><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <i>30-60 calendar days</i> upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance</p>			

