



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2023 – 039E**

### **PROCUREMENT OF ROUTINE IMMUNIZATION CARD**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
Add: 1 spare LED Bulb	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19<sup>th</sup> day of May 2023 in MMCHD

Approved by:

SGD

**PRETCHELL P. TOLENTINO, MD, MCHM,**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p><b>Republic of the Philippines</b></p> <p><b>Department of Health</b></p> <p><b>Metro Manila Center for Health Development</b></p> <p><b>TECHNICAL SPECIFICATIONS</b></p>			
Item No. 1	<b>MICROSCOPE, BINOCULAR</b>	Qty./Unit	5 units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC: 450,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b><u>A. SPECIFICATIONS:</u></b></p> <p>Illumination: Built-in LED illumination system</p> <p>Focusing: - Coaxial coarse and fine adjustments with 200 graduations - Precision of at least 2 microns</p> <p>Nosepiece: Quadruple revolving nosepiece</p> <p>Eyepieces: Wide field 10x/20mm eyepieces</p> <p>Head: - 30° inclined - 360° rotatable - Interpupillary Distance: Adjustable between 55-75 mm</p> <p>Stage: Single specimen holder (1 pc Double specimen holder and 1 pc sheet holder)</p> <p>Condenser: - must be adjustable - achromatic objectives 4x, 10x, 40x and 100x - all anti-fungus treated</p> <p>Power: 220 VAC, 50/60 Hz</p> <p>Accessories: - Power Cord - LED Bulb - Dust Cover - 1 spare LED Bulb</p>			
<p><b><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></b></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The</p>			

