



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 040E
PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
> Mattress: At least 3 cm thick, water proof seamless mattress	> Mattress: At least 2 <i>inches</i> thick, water proof seamless mattress
>• Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 50mm in every dimension, adjustable height must be at least in range or bigger	>• Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 100mm in every dimension, adjustable height must be at least in range or bigger
Delivery: > 30 calendar days.	Delivery: >30-60 calendar days.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25th day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of May 2023 in MMCHD

Approved by:

SGD
PRETCHELL P. TOLENTINO, MD, MCHM,
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Republic of the Philippines</p> <p>Department of Health</p> <p>Metro Manila Center for Health Development</p> <p>TECHNICAL SPECIFICATIONS</p>			
Item No. 1	WHEELED DETACHABLE STRETCHER	Qty./Unit	29 units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 2,030,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>A. SPECIFICATIONS:</u></p> <ul style="list-style-type: none"> • Frame: Aluminum Alloy • Type: Detachable/Removable • Mattress: At least 2 inches thick, water proof seamless mattress • Has 3 belts • Castor wheels: At least 100 mm diameter castors • Max. Load At least 160kg • Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 100mm in every dimension, adjustable height must be at least in range or bigger • Adjustable height: 300mm to 850mm 			
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <i>30-60 calendar days</i> upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for two (2) years on parts and services. The supplier shall replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p>			

