



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2023 – 042E
PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
> Analyzer Configuration: 9 configurations	> Analyzer Configuration: 9 configurations - Na+, K+ - Na+, K+, Cl- - Na+, K+, Ca2+ - Na+, K+, Li+ - Na+, Li+ - Na+, Ca2+, Li+ - Li+ - Na+, Cl- - Na+, Ca2+
Delivery: >30 calendar days.	Delivery: >30-60 calendar days.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25th day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of May 2023 in MMCHD

Approved by:

SGD
PRETCHELL P. TOLENTINO, MD, MCHM,
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	ELECTROLYTES ANALYZER	Qty./Unit	4 units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 1,600,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><u>A. SPECIFICATIONS:</u></p> <ul style="list-style-type: none"> • Analyzer Type: Stand alone, fully-automated • Analyzer Configuration: 9 configurations <ul style="list-style-type: none"> - Na+, K+ - Na+, K+, Cl- - Na+, K+, Ca2+ - Na+, K+, Li+ - Na+, Li+ - Na+, Ca2+, Li+ - Li+ - Na+, Cl- - Na+, Ca2+ • Sample Type: whole blood, plasma, Serum acetate, Bicarbonate Dialysate, aqueous material, prediluted urine • Analysis Time: ≤50 seconds per sample • Sample Size: 95uL • Calibration: <ul style="list-style-type: none"> - auto-calibration every power-on or reset - auto-calibration every measurement - auto-calibration every 4 hours - the analyzer can be programmed to automatically enter standby mode to suspend automatic calibration • Safety Standards: ANSI Z535, IEC-61010-2-101, IEC 60417 and ISO 7000 • Construction: <ul style="list-style-type: none"> - Electrode housing: Transparent Acrylic - Pin connector: silver (Ag), silver chloride (AgCl) - Electrolyte Chamber: airtight with electrolyte solution - Has built-in thermal printer - Thermal printer paper size: 38mm width - Dimensions: Manufacturer's Standard - Display Type: LCD screen • Accessories: 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<ul style="list-style-type: none">- 5 roll thermal printer paper 38mm x 9m- Quad ring: 1.78 x 1.02mm- Fuse set- Sample Probe stylet- 10 mL Syringe with tapered tip- 1 kit of Reagent (At least 300 tests)	
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></p> <ol style="list-style-type: none">1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <i>30-60 calendar days</i> upon receipt of Notice to Proceed.2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:<ol style="list-style-type: none">a) Service manual in English languageb) Operation manual in English language7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit <p>X-X</p> <p>Source of Fund: SAA 2022-02-0531 (HFEP 2022 ConAp)</p> <p>Recipients:</p> <p>Health Facilities in Malabon City – 4 units</p>	