# A DEPART OF HER

### Republic of the Philippines Department of Health

### METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

### IB 2023 – 048E PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

### Revision and clarification to provisions/specifications in the Bidding Documents:

Requirements

Add:

- >3. Training: The Supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.
- >5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- >4. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
- >a) Service Manual in English language
- >b) Operation Manual in English language

r e) of oranion management		
FROM	TO	
>5. With "DOH-MMCHD HFEP" (Government	> 5. With "DOH-MMCHD HFEP" (Government	
Property not for sale) sticker in each	Property not for sale) sticker in each unit.	
container/tray.		
Requirements		
Remove:		
a) Sterilization/Cleaning manual in English language	age	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19<sup>th</sup> day of May 2023 in MMCHD

Approved by:

**SGD** 

PRETCHELL P. TOLENTINO, MD, MCHM,

Director III / BAC Chairperson



### Republic of the Philippines Department of Health

# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### Republic of the Philippines

### **Department of Health**

### Metro Manila Center for Health Development

### TECHNICAL SPECIFICATIONS

Name of Manufacturer:  Brand:  ABC: 320,000.00  PURCHASER'S SPECIFICATION  A. SPECIFICATIONS:  · Material: Strong Plastic Case with Metal Lid  · Brushless motor  · LED Display speed and timer  · Alarm: overspeed and imbalance  · With mechanical door lock and safety switch  · Error code display  · Speed up and speed down adjustment  · Suitable for qualitative analysis of blood serum and plasma  · Max RPM: 4000 RPM or Higher  · Max RCF: at least 2300 x g  · Timer: 0 to 90 minutes	Qty./Unit	4 units
ABC: 320,000.00  PURCHASER'S SPECIFICATION  A. SPECIFICATIONS:  · Material: Strong Plastic Case with Metal Lid  · Brushless motor  · LED Display speed and timer  · Alarm: overspeed and imbalance  · With mechanical door lock and safety switch  · Error code display  · Speed up and speed down adjustment  · Suitable for qualitative analysis of blood serum and plasma  · Max RPM: 4000 RPM or Higher  · Max RCF: at least 2300 x g	Country of Origin	
PURCHASER'S SPECIFICATION  A. SPECIFICATIONS:  Material: Strong Plastic Case with Metal Lid  Brushless motor  LED Display speed and timer  Alarm: overspeed and imbalance  With mechanical door lock and safety switch  Error code display  Speed up and speed down adjustment  Suitable for qualitative analysis of blood serum and plasma  Max RPM: 4000 RPM or Higher  Max RCF: at least 2300 x g	Model: (if applicable)	
A. SPECIFICATIONS:  · Material: Strong Plastic Case with Metal Lid  · Brushless motor  · LED Display speed and timer  · Alarm: overspeed and imbalance  · With mechanical door lock and safety switch  · Error code display  · Speed up and speed down adjustment  · Suitable for qualitative analysis of blood serum and plasma  · Max RPM: 4000 RPM or Higher  · Max RCF: at least 2300 x g		
<ul> <li>Material: Strong Plastic Case with Metal Lid</li> <li>Brushless motor</li> <li>LED Display speed and timer</li> <li>Alarm: overspeed and imbalance</li> <li>With mechanical door lock and safety switch</li> <li>Error code display</li> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>	STATEMENT OF COMPLIANCE	
<ul> <li>Brushless motor</li> <li>LED Display speed and timer</li> <li>Alarm: overspeed and imbalance</li> <li>With mechanical door lock and safety switch</li> <li>Error code display</li> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>LED Display speed and timer</li> <li>Alarm: overspeed and imbalance</li> <li>With mechanical door lock and safety switch</li> <li>Error code display</li> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>Alarm: overspeed and imbalance</li> <li>With mechanical door lock and safety switch</li> <li>Error code display</li> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>With mechanical door lock and safety switch</li> <li>Error code display</li> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>Error code display</li> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>Speed up and speed down adjustment</li> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>Suitable for qualitative analysis of blood serum and plasma</li> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
<ul> <li>Max RPM: 4000 RPM or Higher</li> <li>Max RCF: at least 2300 x g</li> </ul>		
· Max RCF: at least 2300 x g		
5		
· Timer: 0 to 90 minutes		
***************************************		
· Rotor Capacity: 12 by 20 ml/15 ml/10 ml/5 ml		
· Power: 220V, 50/60 Hz		
· Dimension: Manufacturer's Standard		

- 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.
- 2. Prior to acceptance, the end user shall conduct a physical inspection. The instruments must have no physical damage and defect.
- 3. **Training:** The Supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.
- 4. Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

# THE PHILIPPING STATE OF THE PH

### Republic of the Philippines Department of Health

# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- 6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
- a) Service Manual in English language
- b) Operation Manual in English language
- 7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

Source of Fund: SAA 2022-02-0531 (HFEP ConAp 2022)

Recipients:

Marikina City - 2 units

Muntinlupa City - 2 units