



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2023 – 049E**

### PROCUREMENT OF ROUTINE IMMUNIZATION CARD

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
> Caster w/ 4" diameter	>Caster: at least 4" diameter or better.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of May 2023, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19<sup>th</sup> day of May 2023 in MMCHD

Approved by:

SGD

**PRETCHELL P. TOLENTINO, MD, MCHM,**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p><b>Republic of the Philippines</b> <b>Department of Health</b> <b>Metro Manila Center for Health Development</b></p> <p><b>TECHNICAL SPECIFICATIONS</b></p>			
Item No. 1	<b>BED, MECHANICAL, 3 CRANKS WITH MATTRESS</b>	Qty./Unit	7 units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC: 350,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p><b><u>A. SPECIFICATIONS:</u></b></p> <ul style="list-style-type: none"> <li>● 3-crank mechanism</li> <li>● Back and knee raise</li> <li>● Rectangular tubular frame</li> <li>● Sliding full side rails</li> <li>● IV pole &amp; socket</li> <li>● ABS plastic head board and foot board</li> <li>● Caster: at least 4" diameter or <i>better</i>.</li> <li>● Foot lock on all casters (directional)</li> <li>● Leatherette cover mattress</li> <li>● Mattress size: at least 35" width x 4" thick x 76" length</li> <li>● With patient name holder at foot part</li> <li>● Adjustable Height from bed surface is at least 17" - 27"</li> </ul>			
<p><b><u>B. REQUIREMENTS IF AWARDED THE CONTRACT</u></b></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or</p>			

