



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 019E**  
**PROCUREMENT OF FETAL MONITOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
At least 7-inch. TFT color LCD Display	<del>At least 7-inch.</del> TFT color LCD Display or Manufacturer Standard

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 13th day of May 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents that are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of May 2024 in MMCHD

Approved by:

SGD.  
**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson

**IB2024 – 019E PROCUREMENT OF FETAL MONITOR**

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	<b>FETAL MONITOR</b>	Qty./Unit	<b>8 UNITS</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>1,280,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b>TECHNICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>• Dual pulsed Doppler FHR</li> <li>• Ultrasound Frequency: 1 MHz</li> <li>• Fetal Heart Rate Measurement Range: 30-240 bpm</li> <li>• With auto-detection of dual fetal movement</li> <li>• With array-type thermal printer</li> <li>• With automatic printing function</li> <li>• <i>TFT color LCD Display or Manufacturer Standard</i></li> <li>• 3-channel display: FHR I, FHR II, UC</li> <li>• With Clinical event marking</li> <li>• At least 160 patient data saving</li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Can be connected to a central monitoring system through PC interface</li> <li>• Has built-in battery: 4-8 hours</li> <li>• Power aid module: 10-12 hours</li> <li>• Accessories <ul style="list-style-type: none"> <li>- 2 units Ultrasound Doppler probe</li> <li>- 1 unit event marker jack</li> <li>- 3 units probe belt</li> <li>- 1 unit UC probe</li> <li>- 1 unit ultrasound gel</li> <li>- 2 units probe holder</li> </ul> </li> </ul>			
<p><b>Requirements if awarded the Contract</b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. <b>Notarized undertaking</b> that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</p>			

