



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2024-027E  
 PROCUREMENT OF DENTAL UNIT, MOTORIZED WITH ACCESSORIES**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
No changes stipulated in the Technical Specifications	

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 27<sup>th</sup> day of May 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17<sup>th</sup> day of May 2024 in MMCHD

Approved by:

SGD.  
**JEREMIAS FRANCIS Y. CHAN, MD**  
 Licensing Officer V / BAC Chairperson

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

<b>Item No. 1</b>	<b>DENTAL UNIT, MOTORIZED WITH ACCESSORIES</b>	Qty./Unit	<b>16 UNITS</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>2,880,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p><b>TECHNICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>· Fully motorized with pre-set functions</li> <li>· Touch control chair</li> <li>· Slim ergonomic design upholstery</li> <li>· LED light with sensor switch</li> <li>· Double side armrest</li> <li>· Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer)</li> <li>· Two (2) pcs. 3-way syringe</li> <li>· Three (3) pcs handpiece attachments (2 hi-speed adaptors &amp; 1 low-speed adaptor)</li> <li>· Auto-cup-filler &amp; bowl flush (water flushing control valve, cup-filler with control valve)</li> <li>· Saliva ejector-built-in motor pump type, does not use air or water</li> <li>· Clean water system</li> <li>· LED X-ray film viewer (Peri-apical X-ray viewer)</li> <li>· Multi-functional foot control</li> <li>· Water tank for distilled water for handpiece</li> <li>· Two (2) hi-speed handpieces</li> <li>· One (1) low-speed complete with straight handpiece &amp; Contra-angle</li> <li>· LED light cure unit (external)</li> <li>· Tissue box with tray and cup dispenser</li> <li>· Silent-type and oil-free dental air compressor</li> <li>· Saddle-type with backrest dentist stool</li> </ul>			
<p>Requirements if awarded the Contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is</p>			

