



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2024-028E  
 PROCUREMENT OF ECG MACHINE, 12-CHANNEL**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.	<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30 to 60</b> calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 27<sup>th</sup> day of May 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17<sup>th</sup> day of May 2024 in MMCHD

Approved by:

SGD.  
**JEREMIAS FRANCIS Y. CHAN, MD**  
 Licensing Officer V / BAC Chairperson

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

<b>Item No. 1</b>	<b>ECG MACHINE, 12-CHANNEL</b>	Qty./Unit	<b>2 UNITS</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>360,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p><b>TECHNICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>• <b>Features</b></li> <li>- Operates with touch screen and function buttons</li> <li>- Synchronized collection for 12-lead ECG, supports 12-lead and Cabrera-lead waveform</li> <li>- Adopts digital signal processing technology and get high-quality ECG via power frequency filter, baseline filter, EMG filter and Low-pass filter for ECG signals.</li> <li>- Capable of displaying 3/6/12-lead ECG waveform on one screen.</li> <li>- Capable of displaying HR value, print mode, sensitivity, paper speed, filter state, clock, battery level, background grid lines, measured data and interpretation information.</li> <li>- Capable of auto-measurement and auto-interpretation for routine ECG parameters.</li> <li>- Capable of providing measurement results and auto-diagnosis for Heart Rate (HR), PR Interval, P Duration, QRS Duration, T Duration, QT/QTc Interval, P/QRS/T Axis, R(VS), S(VI), R(VS)+S(VI) amplitude, and Cornel Index.</li> <li>- Must be powered by AC and DC</li> <li>• <b>Display:</b> Color LCD with resolution of at least 1280x800</li> <li>• <b>Printer:</b></li> <li>- With a built-in thermal printer that supports auto, manual, and other printing modes.</li> <li>- Printed content must contain time, paper speed, sensitivity, calibration signal, name of lead, filter state, and patients' information</li> <li>• <b>Memory:</b> can store up to 4000 medical records</li> <li>• <b>Recording Paper:</b> high-speed thermal paper</li> <li>• <b>Sampling precision:</b> 24-bit</li> <li>• <b>Waveform data sampling frequency:</b> 1 kHz</li> <li>• <b>Connectivity:</b> Wi-Fi, LAN and USB connection</li> <li>• <b>Battery</b></li> <li>At least 5000 mAh built-in rechargeable battery and charging circuit that can last up to 10 hours during standby and 3 hours during continuous operation.</li> <li>- With battery overcurrent and overvoltage protection circuit.</li> <li>• <b>Power Supply:</b> 220VAC, 50/60Hz</li> </ul>			
<p><b>Requirements if awarded the Contract</b></p> <p>1. <b>Completion Period:</b> <i>The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30 to 60</b> calendar days upon receipt of Notice to Proceed..</i></p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p>			

