



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2024-031E  
PROCUREMENT OF 2 UNITS REFRIGERATOR (LOT BIDDING)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
Power failure backup: 8 hours	Power failure backup: 8 hours (built-in)

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 27<sup>th</sup> day of May 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17<sup>th</sup> day of May 2024 in MMCHD

Approved by:

SGD.  
**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

<b>VACCINE REFRIGERATOR</b>	Qty./Unit	<b>1Unit</b>	
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>250,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b>TECHNICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>• Capacity: at least 80L</li> <li>• Construction <ul style="list-style-type: none"> <li>- Stainless steel Handle</li> <li>- Drop-proof Polyethylene cabinet or better</li> <li>- Silicone seal</li> </ul> </li> <li>• Alarm: <ul style="list-style-type: none"> <li>- Sensor Error Alarm, High/Low temperature alarm, Controller failure alarm</li> <li>- Power failure alarm</li> </ul> </li> <li>• Power Supply: <ul style="list-style-type: none"> <li>- AC: 220V 50/60 Hz</li> <li>- DC: 12V</li> <li>- <i>Power failure backup: 8 hours (built-in)</i></li> </ul> </li> <li>• Refrigeration System: <ul style="list-style-type: none"> <li>- Refrigerant: R134a (CFC-free)</li> <li>- Temperature controller: Microprocessor</li> </ul> </li> <li>• Display: Digital</li> <li>• Temperature Accuracy: 0.1°C</li> <li>• Refrigeration Temperature: 2°C ~ 8°C</li> <li>• Freezer Temperature: -10°C ~ -25°C</li> <li>• Has Data logger included</li> <li>• Noise: 46 dB or better</li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Accessories: <ul style="list-style-type: none"> <li>- 220V power cord, 12V power cord, Plastic Frames</li> </ul> </li> </ul>			
<p><b>Requirements if awarded the Contract</b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. <b>Warranty:</b></p> <p>a) Warranty certificate for two (2) years on parts and service. The</p>			

supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations.

c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects

d) The number of days where the equipment is unusable due to equipment defect/faults shall be added to the warranty period. at least every six (6) month or according to the manufacture's recommendation;

5. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit

x-x

Source of Fund:

SAA 2023-02-000687 (HFEP 2023-CONAP)

Recipient: 1 Unit Health Facilities San Juan City