



SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2024-032E
PROCUREMENT OF 2 UNITS GENERATOR SET, 10 KVA**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
No changes stipulated in the Technical Specifications	

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 27th day of May 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17th day of May 2024 in MMCHD

Approved by:

SGD.
JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	GENERATOR SET, 10KVA	Qty./Unit	4 UNITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 560,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> "● Engine: 4-stroke Diesel Engine" "● Type: Silent" "● Phase: Single" "● Frequency: 60Hz" "● Max AC Output: at least 10kva" "● Rated AC Output: at least 8.5kva" "● Rated AC Voltage: 220V" "● Power factor: 0.8" "● Displacement: 520cc" "● Fuel: Diesel Light Fuel" "● Fuel Tank Capacity: at least 16L" "● Ignition system: Direct injection" "● Operating Noise Level: 75dB" "● Dimensions: Manufacturer's Standard" 			
<p>Requirements if awarded the Contract</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period. 6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following: <ol style="list-style-type: none"> a) Service manual in English language b) Operation manual in English language 7. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in 			

