



Republic of the Philippines  
 Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 034E**  
**PROCUREMENT OF DENTAL UNIT, MOTORIZED WITH ACCESSORIES**  
**(REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
<b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.	

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 14<sup>th</sup> day of June 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7<sup>th</sup> day of June 2024 in MMCHD

Approved by:

SGD.  
**JEREMIAS FRANCIS Y. CHAN, MD**  
 Licensing Officer V / BAC Chairperson

## VII Technical Specification

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
<b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>DENTAL UNIT, MOTORIZED WITH ACCESSORIES</b>	Qty./Unit	<b>4 UNITS</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>720,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>TECHNICAL SPECIFICATIONS:</b> <ul style="list-style-type: none"> <li>· Fully motorized with pre-set functions</li> <li>· Touch control chair</li> <li>· Slim ergonomic design upholstery</li> <li>· LED light with sensor switch</li> <li>· Double side armrest</li> <li>· Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer)</li> <li>· Two (2) pcs. 3 way syringe</li> <li>· Three (3) pcs hand piece attachments (2 hi-speed adaptors &amp; 1 low-speed adaptor)</li> <li>· Auto-cup-filler &amp; bowl flush (water flushing control valve, cup-filler with control valve)</li> <li>· Saliva ejector-built-in motor pump type, does not use air or water</li> <li>· Clean water system</li> <li>· LED X-ray film viewer (Peri-apical X-ray viewer)</li> <li>· Multi-functional foot control</li> <li>· Water tank for distilled water for hand piece</li> <li>· Two (2) hi-speed hand piece</li> <li>· One (1) low-speed complete with straight hand piece &amp; Contra-angle</li> <li>· LED light cure unit (external)</li> <li>· Tissue box with tray and cup dispenser</li> <li>· Silent type and oil-free dental air compressor</li> <li>· Saddle-type with backrest dentist stool</li> </ul>			
<b>Requirements if awarded the Contract</b> <p>1. <b>Completion Period:</b> <i>The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed..</i></p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship</p>			

