



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB2024 – 035E
PROCUREMENT OF DENTAL CHAIR, PEDIATRIC (REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.	

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 14th day of June 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7th day of June 2024 in MMCHD

Approved by:

SGD.
JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

VII Technical Specification

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	DENTAL CHAIR, PEDIATRIC (REBID)	Qty./Unit	4 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,200,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Special design for children • All plastics are made of injection mold • Normal operating light • Normal suction system • Three-way syringe (One for hot and one for cold), 2 pcs • Built-in water purified system • Automatic water heating system • Built-in X-ray film viewer • With 2 pieces of dental stool (1 saddle type) • Inclusions <ul style="list-style-type: none"> - 1 unit of Intra Oral Camera with at least 17" LED monitor - 1 unit of Ultrasonic Scaler with LED (with 4 tips included) - 1 unit of Light Cure Machine - 3 pieces of Handpiece 4 holes or 2 holes - 2 hp Compressor with at least 25L capacity and 115 psi of max. pressure 			
<p>Requirements if awarded the Contract</p> <p>1. Completion Period: <i>The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed...</i></p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital</p>			

