

## Republic of the Philippines Department of Health

# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

#### SUPPLEMENTAL/ BID BULLETIN NO. 1

### IB2024 – 043E PROCUREMENT OF 26 UNITS TRANSFER STRETCHER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Dimensions (L x W x H): 2000mm x 700mm x 700mm, ± 100mm on each dimension	Dimensions (L x W <del>x H</del> ): 2000mm x 700mm <del>x 700mm</del> , ± 100mm on each dimension	Dimensions (L x W): 2000mm x 700mm , ± 100mm on each dimension
Back raise inclination: up to 65°	Back raise inclination: up to 65° or better	Back raise inclination: up to 65° or better
Adjustable height: At least 600 - 900 mm	Adjustable height: At least 550 or 600 – 850 or 900 mm	Adjustable height: : At least 550 or 600 – 850 or 900 mm
Equipped with two-stage IV Stand, adjustable up to 100 cm	Equipped with two-stage IV Stand—Pole adjustable up to 100 cm	Equipped with a two-stage IV Pole adjustable up to 100 cm

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 14<sup>th</sup> day of June 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7th day of June 2024 in MMCHD

Approved by:

SGD.

JEREMIAS FRANCIS Y. CHAN, MD

Licensing Officer V / BAC Chairperson

۷.

#### **VII Technical Specification**

### Republic of the Philippines Department of Health Metro Manila Center for Health Development **TECHNICAL SPECIFICATIONS** Item No. 1 **PROCUREMENT OF** Qty./Unit **26 UNITS** TRANSFER STRETCHER Name of Manufacturer: Country of Origin Brand: Model: (if applicable) ABC: **1,820,000.00** PURCHASER'S SPECIFICATION STATEMENT OF COMPLIANCE **TECHNICAL SPECIFICATIONS:** • Dimensions (L x W): 2000mm x 700mm, ± 100mm on each dimension • Material: made of powder coated steel and ABS plastic • Lifting mechanism: Manual • **Back raise inclination**: up to 65° or better • Adjustable height: At least 550 or 600 - 850 or 900 mm • Equipped with side guard rails • Equipped with two-stage IV Pole, adjustable up to 100 cm • Equipped with a central control brake device • Castor wheel size: Approx. 150 mm in diameter Accessories: - Has oxygen cylinder holder - Has IV pole holder located in 4 different positions - With one IV pole Requirements if awarded the Contract 1. **Completion Period**: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed. 2. **Testing**: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3.**Training**: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.

4. Warranty

a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by

the end-user after testing and commissioning.

- b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;
- c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.
- d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.
- e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.
- 5. **Notarized** undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- 6. **Manuals**: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
  - a) Service manual in English language
  - b) Operation manual in English language
- 7. With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit

Recipient: Caloocan city Medical Center