

## Republic of the Philippines Department of Health METRO MANILA CENTER FOR HEALTH DEVELOPMENT

#### SUPPLEMENTAL/ BID BULLETIN NO. 1

### IB2024 – 044E PROCUREMENT OF RADIANT WARMER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:				
Technical Specifications	Query	Response of the End User Unit		
Ceramic heater elements and golden reflectors.	Ceramic heater elements and golden reflectors or its equivalent	Ceramic heater elements and golden reflectors or <i>its</i> <i>equivalent</i>		
With at least 10 heater level.	With at least 5 heater level or better	With at least 5 heater level or better		
With at least 2 heater	With at least 1 heater	With at least 1 heater		
elements located at the	elements located at the	elements located at the		
Radiant warmer	Radiant warmer or better	Radiant warmer <i>or better</i>		
Manual bed-tilt feature Head	Manual bed-tilt feature Head	Manual bed-tilt feature Head		
up position is 20 degrees.	up position is 15 degrees or	up position is 15 degrees or		
	better	better		
- Manual bed-tilt feature Head	Manual bed-tilt feature Head	Manual bed-tilt feature Head		
down position is 15 degrees	down position is 12 degrees or better	down position is <i>12 degrees or better</i>		

#### 2. Completion Period:

The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit them together with all required documents for the submission of bids on the 14<sup>th</sup> day of June 2024, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7<sup>th</sup> day of June 2024 in MMCHD

Approved by:

SGD. JEREMIAS FRANCIS Y. CHAN, MD Licensing Officer V / BAC Chairperson

# VII Technical Specification

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
Item No. <b>1</b>	TECHNICAL SPECIFICATI PROCUREMENT OF RADIANT WARMER	Qty./Unit	5 UNITS
Name of Manufac		Qty./Onit	Country of Origin
		Country of origin	
Brand:			Model: (if applicable)
ABC: 1,250,000.	00		
	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE
<b>TECHNICAL SPE</b>			
Movable Lamp	mechanism that automatically focuses radian	t warmth on	
the baby.			
• Lamp can be	• Lamp can be positioned on the overhead left or right side of the		
equipment.			
	• Ceramic heater elements and golden reflectors <i>or its equivalent</i>		
	Radiant warmth with Two (2) brightness level Examination lamp features		
	heater level or better		
• Radiant Warmer can be positioned from left to right direction for up to			
90 degrees.			
• With at least <i>1 heater</i> elements located at the Radiant warmer <i>or better</i>			
• With 30 watts day and 9 watts night examination light.			
• Radiant Power at Heat Level 3 between the bed and radiant warmer: 10mW/cm2			
• Radiant Power at Heat Level 10 between the bed and radiant warmer:			
30mW/cm2			
• Lamps (Work Light): 30 Watts			
	• Lamps (Night Light): 9 Watts		
• Minimum clearance between top edge of Radiant warmer and			
Ceiling: >50cm			
• Temperature			
- Capable of measuring baby's central or peripheral temperature.			
- Manual mode			
- Skin servo mode			
- With temperature measuring range of at least 15 degrees to 42 degrees			
<ul> <li>Accuracy (Sensor): ±0.1 degrees Celsius</li> <li>Setting Range: 30 degrees Celsius to 38.5 degrees Celsius</li> </ul>			

Variable Height Adjustment (VHA): At least 290mm	
Standard Features	
- Swivel cupboards located at right or left side of the machine.	
- Manual bed-tilt feature Head up position is 15 degrees or better	
- Manual bed-tilt feature Head down position is 12 degrees or better	
- The inner side walls must be 7cm.	
- Equipped with Standard X-ray tray.	
Accessories	
- One (1) unit Skin temperature probe	
- One (1) unit Power cable	
- One (1) unit Power adapter	
Requirements if awarded the Contract	
1. Completion Period: The delivery, installation, testing and	1
commissioning of the equipment and its accessories, including the training	
of end-users and maintenance staff must be completed with 60 calendar	
days upon receipt of Notice to Proceed.	
<b>2. Testing</b> : Prior to acceptance, the end user shall conduct a physical	
inspection and functionality test. The equipment must be functioning and	
must have no physical damage and defect.	
3.Training: The supplier shall provide a training on the proper use and	
maintenance of the equipment to the end-users and to the hospital	
maintenance staff within 3 days upon the delivery of the equipment.	
4. Warranty	
a) Warranty certificate for two (2) years on parts and service. The	
supplier shall either repair or replace any item or part in the equipment	
that is found to be defective in material or workmanship under normal	
use. The warranty period shall commence from the date of acceptance by	
the end-user after testing and commissioning.	
b) Preventive maintenance at least every six (6) months or according to	
the manufacturer's recommendations;	
c) Corrective maintenance within five (5) calendar days upon	
notification from the end-user regarding equipment breakdown/defects.	
d) The number of days where the equipment is unusable due to	
equipment defects/faults shall be added to the warranty period.	
e) The supplier shall specify post-warranty comprehensive preventive	
maintenance costs including list and prices of major spare parts of the	
equipment for three (3) years after the warranty period.	
5. Notarized undertaking that the supplier shall conduct the necessary	
corrective maintenance within five (5) calendar days upon notification of	
the equipment breakdown from the end-user. The undertaking shall	
include a statement that the number of days where the equipment is	
unusable due to defective material or workmanship, shall be added to the	
warranty period.	
6. <b>Manuals</b> : The supplier must provide the end-user one (1) hard and one	
(1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With " <b>DOH-MMCHD HFEP</b> "(Government Property not for sale) sticker	
in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-	
Source of Fund: SAA 2023-02-000687 (HFEP 2023 ConAp)	
Recipient: Caloocan City Medical Center	