



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB2024 – 045E
PROCUREMENT OF MORTUARY REFRIGERATOR (2-BODIES)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
To include:		
End Opening		

2. Completion Period:

The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 14th day of June 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7th day of June 2024 in MMCHD

Approved by:

SGD.
JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

VII Technical Specification

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PROCUREMENT OF MORTUARY REFRIGERATOR (2-BODIES)	Qty./Unit	1 UNIT
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 650,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Rooms: 2 or better • Volume: at least 3.1 cubic meters or better • Material: Stainless steel 304 • Power supply: 220V, 50/60hz • Temperature: at least 0-10°C • Refrigeration: R404A or R290 or better • Must be CE certified • Accessories: Standard manufacturer's accessories <p><i>To include:End Opening</i></p>			
<p>Requirements if awarded the Contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty</p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</p>			

