



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 046E

PROCUREMENT OF LABORATORY INCUBATOR

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
No changes stipulated in the Technical Specifications		

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 14th day of June 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7th day of June 2024 in MMCHD

Approved by:

SGD.
JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

VII Technical Specification

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PROCUREMENT OF LABORATORY INCUBATOR	Qty./Unit	2 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 600,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Capacity: At least 50 L chamber capacity • Dimensions: Manufacturer's Standard <p>CONSTRUCTION</p> <ul style="list-style-type: none"> • The incubator's main body shall be constructed from 1.2mm 18-gauge electro-galvanized steel • The incubator shall come with antimicrobial coating on all painted surfaces to inhibit microbial growth • The incubator chamber is 1.2mm 18-gauge stainless steel 304 • The incubator shall have a 2-point door seal and eccentric hinge to ensure maximum gasket compression • Fan: must be permanently lubricated and maintenance-free • The incubator shall come with an access port for temperature validation and mapping • The incubator shall have an inner glass door for sample viewing • Power: 220VAC, 50/60Hz • The incubator must employ forced air convection system • Fan can be set to 0 to emply the natural convection system • Fan speed shall be adjustable for optimal air exchange rates • With multiple, adjustable over-temperature protection system • Temperature variation: at 37°C is ±0.3°C • Should have a microprocessor PID control technology for all incubator functions • Connected to an instrument-grade precision platinum temperature probe • Ensures fast ramp time • Prevents overshoot • Ensures stable temperature once set point is achieved • Has twin temperature display (Actual and Set Point) • Audible and visual alarms for over-temperature 			
<p>Requirements if awarded the Contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a</p>			

physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.

4. Warranty

a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;

c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.

d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.

e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit

x-x

Source of Fund: SAA 2023-02-000687 (HFEP 2023 ConAp)

Recipient: Navotas City Hospital