



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB2024 – 048E
PROCUREMENT OF AUTOMATED EXTERNAL DEFIBRILLATOR (AED)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Defibrillator Pads: self-adhesive, disposable and non-polarized - at least five (5) pairs for adult - at least five (5) pairs for pedia	Defibrillator Pads: self-adhesive, disposable and non-polarized or polarized – for clarification	Defibrillator Pads: self-adhesive, disposable and non-polarized - RETAIN
Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed		

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 14th day of June 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 7th day of June 2024 in MMCHD

Approved by:

SDG.
JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

VII Technical Specification

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	PROCUREMENT OF AUTOMATED EXTERNAL DEFIBRILLATOR (AED)	Qty./Unit	2 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 300,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> • Two-button Operation: On/Off, Discharge (shock button) • Voice Prompts: Voice Command that gives instruction to the user in operating the device • Automatically evaluates patient impedance for proper pad contact. • Automatically analyses patient condition for shockable or non-shockable rhythms • Patient ECG can be acquired through defibrillator pads • Energy output accuracy: $\pm 15\%$ • Pulse shape: Biphasic • Charge time: maximum of 15 seconds • Battery Capacity: Capable of providing at least 200 discharges at maximum energy • Replaceable battery • Visible and Audible Indicators: <ul style="list-style-type: none"> - Low battery - Pad skin contact/disconnection - AED Status • Defibrillator Pads: self-adhesive, disposable and non-polarized <ul style="list-style-type: none"> - at least five (5) pairs for adult - at least five (5) pairs for pedia 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty 			

