



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 069

PROCUREMENT OF 12,500 TESTS OF COVID ANTIGEN RAPID TEST KIT

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
PROCUREMENT TITLE: PROCUREMENT OF 12,500 KITS OF COVID ANTIGEN RAPID TEST KIT	PROCUREMENT TITLE: PROCUREMENT OF 12,500 TESTS OF COVID ANTIGEN RAPID TEST KIT
Unit: Kit	Unit: Test


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	PROCUREMENT OF 12,500 TEST OF COVID ANTIGEN RAPID TEST KIT	Qty./Unit	12,500 TESTS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: P 5,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. <u>Specifications:</u></p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> *Specificity: 90-98% *Sensitivity: \geq 90% *Testing Method: Nasopharyngeal Swabbing (Point of Care) *Test Time: 15-20 minutes *Box content: <ul style="list-style-type: none"> *25 Test Devices individually pouched with dessicant, 25 Extraction Tubes, 25 Extraction Cap, *1 kit with positive and negative controls *25 Sterilized Nasopharyngeal Swabs for Sample Collection *1 Sample Tube Rack with self-contained tube with break-off swab *1 Quick Reference Guide, 1 Insertion for Use *Valid FDA Certification *DOH Recommended *Correlation study with RT-PCR Test Internal of Third Party including but not limited to RITM, NIH, FDA, WHO 			
<p>B. <u>Upon delivery the following must be complied:</u></p> <p>Shelf Life:</p> <p>Supplies must be fresh commercial stock with a total shelf life of at least fifteen (15) months from the date of delivery</p> <p>Packaging Instruction:</p> <p>Twenty-five (25) test devices in a box. Number of box per carton will follow standard packaging of manufacturer.</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Transport and Storage Requirement

The storage and working temperature is between 2-30°C

Labelling Instructions:

*Each box should be imprinted or stickered with non-removable or permanent sticker or label that is binding and with residue and tearing if removed:

Philippine Government Property-Department of Health-MMCHD

Not For Sale

Date of Manufacture: _____

Date of Expiry: _____

Batch/ Lot No. _____

Delivery period: 30 days after receipt of PO

Place of Delivery: MMCHD Pasig Warehouse

C. Additional requirement to be submitted by the Single/ Lowest Calculated Bidder (SCB/ LCB) as part of post-qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.