



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 070E**

**PROCUREMENT OF VARIOUS DENTAL EQUIPMENT (LOT BIDDING)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Item No. 1; 2 UNITS DENTAL UNIT, MOTORIZED WITH ACCESSORIES *No changes as stated in the technical specifications.	
Item No. 2: 1 UNIT DENTAL AEROSOL SUCTION MACHINE *No changes as stated in the technical specifications.	
<b>FROM</b>	<b>TO</b>
Item No. 3: 1 UNIT DENTAL ULTRASONIC SCALER MACHINE Recipient: Ospital ng Tondo	Recipient: Valenzuela City Emergency Hospital
<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Item No. 4: 5 UNITS STERILIZING UNIT *No changes as stated in the technical specifications.	
<b>FROM</b>	<b>TO</b>
Item No. 5: 1 SET DENTAL INSTRUMENT SET - Aspirating Syringe, stainless steel, with locking mechanism	Aspirating Syringe, stainless steel, with locking mechanism CW Type
- Minnesota Retractor Stainless Steel	Minnesota Tongue and cheek Retractor Stainless Steel
- Bone Curette, non-magnetic hollow handle	Bone Curette, non-magnetic hollow handle (reunder or oval, 3")
- Gum Separator, non-magnetic hollow handle	- Gum Separator, non-magnetic hollow handle, KPT2
- Hemostat Mosquito Forceps, stainless steel	Hemostat Mosquito Forceps, stainless steel straight
- Perisoteal Elevator, stainless steel	Periosteal Elevator, stainless steel (double edge)
- Bone Rongeur, stainless steel	Bone Rongeur, stainless steel, side cutting
- Mouth Mirror, stainless steel	Mouth Mirror, stainless steel with handle #5 stainless steel
- Explorer, stainless steel	Explorer, stainless steel hollow handle
- Gracey Curette, set of 6, different tips, non-magnetic hollow handle	- Gracey Curette, set of 7, different tips, non-magnetic hollow handle

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23<sup>rd</sup> day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of May 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli", is written over the printed name.

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

## Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>DENTAL UNIT, MOTORIZED WITH ACCESSORIES</b>	Qty./Unit	<b>2 Units</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>360,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>Technical Specifications:</b> DENTAL UNIT, MOTORIZED WITH ACCESSORIES <ul style="list-style-type: none"> <li>· Fully motorized with pre-set functions</li> <li>· Touch control chair</li> <li>· Slim ergonomic design upholstery</li> <li>· LED light with sensor switch</li> <li>· Double side armrest</li> <li>· Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer)</li> <li>· Two (2) pcs. 3 way syringe</li> <li>· Three (3) pcs hand piece attachments (2 hi-speed adaptors &amp; 1 low-speed adaptor)</li> <li>· Auto-cup-filler &amp; bowl flush (water flushing control valve, cup-filler with control valve)</li> <li>· Saliva ejector-built-in motor pump type, does not use air or water</li> <li>· Clean water system</li> <li>· LED X-ray film viewer (Peri-apical X-ray viewer)</li> <li>· Multi-functional foot control</li> <li>· Water tank for distilled water for hand piece</li> <li>· Two (2) hi-speed hand piece</li> <li>· One (1) low-speed complete with straight hand piece &amp; Contra-angle</li> <li>· LED light cure unit (external)</li> <li>· Tissue box with tray and cup dispenser</li> </ul>			





Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	
---	--



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No. 2</b>	<b>DENTAL AEROSOL SUCTION MACHINE</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 60,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>Technical Specifications:</b> <ul style="list-style-type: none"> <li>● Super Bio Filter: 0.3 microns HEPA Filter</li> <li>● With UV light inside the chamber</li> <li>● Filter change alarm function</li> <li>● Noise: ≤ 65 dB</li> <li>● Digital display controls</li> <li>● Triple Filter System               <ul style="list-style-type: none"> <li>- Aerosol Filter</li> <li>- Moisture Contaminant and Dust Filter</li> <li>- Clean Air Filter</li> </ul> </li> <li>● Configurations               <ul style="list-style-type: none"> <li>- Powder-coated steel</li> <li>- Airflow: At least 4000 L/min</li> <li>- Suction Pressure: 25 kpa</li> <li>- Power Supply: 220-240 VAC, 50/60 Hz</li> </ul> </li> </ul>			
<b><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></b> <ol style="list-style-type: none"> <li>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</li> <li>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. <b>Training:</b> The supplier shall provide a training on the proper use and</li> </ol>			



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.

**4. Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

**6. Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

x-x

*Recipient: Valenzuela City Emergency Hospital*

**Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:**


1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 3	<b>DENTAL ULTRASONIC SCALER MACHINE</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>28,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> <li>● Detachable Hand piece with Light (LED), better visibility, autoclavable"</li> <li>● Automatic Frequency Tracking, scaling, perio, endo &amp; clinical application</li> <li>● Automatic Water Supply System with Self - Contained Clean water system</li> <li>● Heat free operation</li> <li>● Ten (10) working tips, includes Perio, Scaler and Endodontic</li> <li>● Endodontic Tip attachment</li> <li>● Power Input: 220 - 240V / 50-60Hz</li> <li>● Output Power: 3W - 20W</li> <li>● Frequency: 28kHz ± 3kHz</li> <li>● Main Unit Weight: Manufacturer's Standard</li> <li>● Adapter Weight: Manufacturer's Standard</li> <li>● Dimension: Manufacturer's Standard</li> </ul>			
<b>REQUIREMENTS IF AWARDED THE CONTRACT:</b>  1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.  2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.  3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.			





## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

**4. Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

**6. Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Service manual in English language

b) Operation manual in English language

7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

x-x

*Recipient: Valenzuela City Emergency Hospital*

**Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No. 4</b>	<b>STERILIZING UNIT</b>	Qty./Unit	<b>5 Units</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>87,500.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<b>Technical Specifications:</b> 2-in-1 Dry Heat/Ultraviolet Sterilizer Power Supply: 220V, 50-60Hz Rated Power: 500W Loading Limit: at least 5 kg Sterilizing Temperature: 40°C, 125°C Dimensions: Manufacturer's Standard Stainless Steel Body Frame Three-Layer Tray each Side Build Timer: 30-60 mins (Dry Heat) Can be wall-mounted Splash Proof Heater Bulb			
<b><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></b>  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.  <b>2. Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.  <b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.  <b>4. Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in			



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit.

x-x

Recipient: Ospital ng Tondo

**C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
<b>TECHNICAL SPECIFICATIONS</b>			
Item No. 5	<b>DENTAL INSTRUMENT SET</b>	Qty./Unit	<b>1 Set</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>45,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>DENTAL INSTRUMENT SET</b> Technical Specifications: Each instrument must be packed individually and labeled by the manufacturer One Set must contain ALL of the following instruments: <b>DENTAL SURGERY SET</b> - Extraction Forceps #16, with cross serration for better grip, ergonomic for better comfort designed instrument - Extraction Forceps #17 with cross serration - Extraction Forceps #18L with cross serration - Extraction Forceps #18R with cross serration - Extraction Forceps #44 with cross serration - Extraction Forceps #69 with cross serration - Extraction Forceps #150 with cross serration - Extraction Forceps #151 with cross serration Aspirating Syringe, stainless steel, with locking mechanism CW Type Minnesota Tongue and cheek Retractor Stainless Steel - Bone File, stainless steel - Bone Curette, non-magnetic hollow handle (reunder or oval, 3") - Gum Separator, non-magnetic hollow handle, KPT2 - Seldin, Stainless steel - Root Tip Pick, straight, non-magnetic hollow handle - Root Tip Pick, left, non-magnetic hollow handle - Root Tip Pick, right, non-magnetic hollow handle			



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- Blade Handle #3, stainless steel
- Elevator #301, non-magnetic hollow handle
- Elevator #304, non-magnetic hollow handle
- Surgical Mallet, stainless steel
- Iris, straight, stainless steel
- Iris, curved, stainless steel
- Adson Tissue Forceps, stainless steel
- Molt #9, stainless steel
- Hemostat Mosquito Forceps, stainless steel straight
- Periosteal Elevator, stainless steel (double edge)
- Bone Rongeur, stainless steel, side cutting
- Allis Tissue, stainless steel
- Needle Holder, stainless steel
- Apical elevator, right, non-magnetic hollow handle
- Apical elevator, left, non-magnetic hollow handle
- Cryer, east, non-magnetic hollow handle
- Cryer, west, non-magnetic hollow handle
- Pedo Forceps, 150SK, stainless steel
- Pedo Forceps, 151SK, stainless steel
- Pedo Forceps, 17S, stainless steel
- Pedo Forceps, 16S, stainless steel
- Pedo Forceps, 18R, stainless steel
- Pedo Forceps, 18L, stainless steel
- Instrument Cassette for sterilization, large, with locking mechanism
  
- Mouth Mirror, stainless steel with handle #5 stainless steel
- Cotton Plier with Lock, stainless steel
- Explorer, stainless steel hollow handle
- Spoon excavator, stainless steel

### **DENTAL PROPHYLAXIS INSTRUMENT SET**

- Universal Scaler, non-magnetic hollow handle
- Perio Curette, non-magnetic hollow handle
- Periodontal Probe
- Gracey Curette, set of 7, different tips, non-magnetic hollow handle



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p><b>RESTORATIVE INSTRUMENTS</b></p> <ul style="list-style-type: none"> <li>- Woodson, non-magnetic hollow handle</li> <li>- Ash, stainless steel</li> <li>- Cement Spatula, stainless steel</li> <li>- Articulating paper forceps, straight, stainless steel</li> <li>- Ball Burnisher, stainless steel</li> <li>- Anatomic Burnisher, stainless steel</li> <li>- Posterior Occlusal Contouring Instrument, stainless steel</li> <li>- Anterior Composite Filling Instrument, stainless steel</li> <li>- Matrix Band Holder, stainless steel</li> <li>- Composite Burnisher, stainless steel</li> <li>- Plugger, stainless steel</li> <li>- Mouth Mirror with Handle #5, stainless steel</li> <li>- Cotton Plier with Lock, stainless steel</li> <li>- Explorer, stainless steel</li> <li>- Spoon Excavator, stainless</li> <li>- Plastic Filling Instrument, stainless steel</li> <li>- Set of Burs, Round-10 Diamond</li> <li>- Laser mark "DOH - MMCHD" in each instrument.</li> </ul>	
<p><b><u>REQUIREMENTS IF AWARDED THE CONTRACT</u></b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Warranty:</b> Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> <li>a) Service manual in English language / Sterilization manual in English</li> </ul> <p>x-x</p> <p>Recipients: Valenzuela City Emergency Hospital</p>	



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

**Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.