



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 071E

PROCUREMENT OF 12 UNITS BED, PLAIN WITH MATTRESS AND IV STAND

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Additional: Provision for side table	
FROM	TO
ABC: P 416,000.00	ABC: P 416,400.00
• Dimensions: Manufacturer's Standard	• Dimensions: At least 2030 x 900 x 500 mm
• Side rails: at least ½ length	• Side rails: at least ½ length (left and right)
• Cover: Leatherette cover with DOH letters printed in the center (front)	• Cover: Leatherette cover with DOH letters printed in the center (front), color green, 6'' tall

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23rd day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15th day of May 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	BED, PLAIN WITH MATTRESS AND IV STAND	Qty./Unit	12 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 416,400.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Dimensions: At least 2030 x 900 x 500 mm • Material: Aluminum alloy or better • Side rails: at least ½ length (left and right) • Mattress: 4 in thick • Cover: Leatherette cover with DOH letters printed in the center (front), color green, 6" tall • Mounted on protective stumps – provision for IV rod (IV socket); 4 (one on each side) • With patient name holder at foot and head part" • Knock-down construction Additional: Provision for side table			
<u>REQUIREMENTS IF AWARDED THE CONTRACT:</u> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 			

