



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 084E

PROCUREMENT OF VARIOUS MEDICAL EQUIPMENT (LINE BIDDING) (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Project Title: PROCUREMENT OF VARIOUS MEDICAL EQUIPMENT (LOT BIDDING) (REBID)	Project Title: PROCUREMENT OF VARIOUS MEDICAL EQUIPMENT (LINE BIDDING) (REBID)
Item No. 1: 4 UNITS INFUSION PUMP Display: At least 4 inches, LCD with big font	Display: At least 3.5 inches, LCD or better
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 2: 6 UNITS NEBULIZER HEAVY DUTY *No changes as stipulated in technical specifications	
FROM	TO
Item No. 3: 4 UNITS ECG MACHINE Paper size: At least 80 mm x 20 mm, roll type	Paper size: At least 50 mm x 20 m, roll type
Recipient: Health Facilities in Paranaque City	Recipient: Health Facilities in Paranaque City – 2 units Ospital ng Paranaque I – 1 unit Ospital ng Paranaque II – 1 unit
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 4: 3 SETS SUCTION MACHINE Type: Mobile with wheels for transport Power supply: 220V, 60Hz Material: Glass or better	
FROM	TO
Item No. 5: 1 UNIT OPERATING ROOM LIGHT 4 BULB (LED) MOBILE (OR LIGHT) Bulb voltage: AC24V Bulb power	Bulb voltage: Manufacturer's standard
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 6: 2 UNITS ELECTRO CAUTERY MACHINE with cart with wheels	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23rd day of May 2022, 9:00 AM:



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.


For guidance and information of all concerned.

Issued this 30th day of May 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "A. Sudiacal", is written over the printed name.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	INFUSION PUMP	Qty./Unit	4 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 340,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> • Horizontal Door Clip • Display: At least 3.5 inches, LCD or better • Ergonomics operation interface • Modes <ul style="list-style-type: none"> - Rate Mode - Drop Mode - Time Mode • Micro Infusion Speed: 0.1ml/h up to 1500 ml/h • Infusion Accuracy: <math>\pm 3\%</math> • With software adjusting function suitable for any brand of infusion sets • With high precision pressure sensor • With built-in drip sensor • Capable of storing up to 800 infusion records and configuration memory • Power Supply: Built-in battery (manufacturer's standard), can last up to 8 hours of operation if fully charged 			
<p><u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Service manual in English language

b) Operation manual in English language


7. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit.

X-X

Recipient: Ospital ng Paranaque 1 - 4 units


c. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 2	NEBULIZER HEAVY DUTY	Qty./Unit	6 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 108,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • Dimensions: Manufacturer's Standard • Compressor Pressure: At least 30 psig • Compressor Type: Diaphragm • Free Air Flow: At least 9 lpm • Operating Pressure: 12-18 psi • Sound Level: ≤ 55 dBA • Power Supply: 220V, 60Hz • Inclusions <ul style="list-style-type: none"> - Mouthpiece - T-piece - Corrugated Tubing - Baffle - Medication Cup - Nebulizer Air-Inlet Connector - Pacifier - Adult/Pedia Aerosol mask - Compressor Filter - Filter Kit (6 each - caps and filters) - Tubing Connector (package of 6) 			
<u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT


Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 3	ECG MACHINE	Qty./Unit	4 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 280,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> • Paper size: At least 50 mm x 20 m, roll type • Automatic measurement & interpretation • Manual/Auto/Arrhythmia Analysis mode or equivalent • Rhythm lead with histogram & trend graph for RR interval or equivalent • With graphic LCD to display 1 channel ECG waveform, setting menu and operating conditions of the unit" • Dual operation, AC & Li-ion battery supporting 2 hours of operation 			
<p><u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

6. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

X-X

Recipient:


Health Facilities in Paranaque City – 2 units

Ospital ng Paranaque I – 1 unit

Ospital ng Paranaque II – 1 unit

C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 4	SUCTION MACHINE	Qty./Unit	3 Sets
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 255,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> • Noise, not more than ≤ 61 dB • Pumping rate: At least 20 L/min • Reservoir capacity: 1000 ml/pe • Reservoir: 2 • Adopt oil free-high-quality pump • Type: Mobile with wheels for transport • Power supply: 220V, 60Hz • Material: Glass or better 			
<p><u>B Requirements if awarded the Contract</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff.</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to</p>			




Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 5	OPERATING ROOM LIGHT 4 BULB (LED) MOBILE (OR LIGHT)	Qty./Unit	1 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 40,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> • Illumination: $\geq 50,000\text{LX}$ (at 1m distance) • Color temperature: $4000\text{K} \pm 500\text{K}$ • Bulb voltage: Manufacturer's standard • Power Supply: 220V, 60Hz • Inclusion • - UPS 			
<p><u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall 			



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

C Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.