



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 085E**

**PROCUREMENT OF VARIOUS MEDICAL FURNITURE (LINE BIDDING) (REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Item No. 1: 2 UNITS DENTAL INSTRUMENT CABINET Stainless steel grade 304	
FROM	TO
Size: Manufacturer's Standard	Size: at least HxLxD 820mm x 500mm x 500mm
Item No. 2: 6 UNITS INSTRUMENT CABINET Stainless steel sheet with two side and back stainless	
Three Glass Shelves (adjustable)	Three Stainless steel shelves (adjustable)
Base compartment with two stainless steel doors, one shelf with lock mounted on heavy-duty rubber casters with lock	<ul style="list-style-type: none"> <li>• Base compartment with two stainless steel doors, one shelf, with lock</li> <li>• Mounted on heavy-duty rubber casters with lock</li> </ul>
Dimensions: Manufacturer's Standard	Dimension: at least HxLxD 1800mm x 800mm x 400mm
<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Item No. 2: 6 UNITS INSTRUMENT CABINET Stainless steel grade 304	
Item No. 3: 5 UNITS INFANT WEIGHING SCALE, DIGITAL *No changes as stipulated in technical specifications	
Item No. 4: 7 UNITS MEDICINE CABINET Stainless steel grade 304	
FROM	TO
Item No. 4: 7 UNITS MEDICINE CABINET Stainless steel sheet with two side stainless steel	Material: Stainless steel 304
Two glass doors with lock	Two glass doors with lock (Type of glass: Manufacturer's Standard)
Three glass shelves (adjustable) or three stainless steel shelves (adjustable)	three stainless steel shelves (adjustable)
<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Item No. 5: 2 UNITS ER BED / TRANSFER STRETCHER *No changes as stipulated in technical specifications	
Item No. 6: 8 UNITS DELIVERY TABLE Stainless steel grade 304 Stool, Stainless Steel (Height adjustable)	
FROM	TO
Item No. 6: 8 UNITS DELIVERY TABLE - 1 pc dirt bucket	- 1 pc dirt bucket stainless steel with wheels
Item No. 7: 2 UNITS WEIGHING SCALE WITH HEIGHT MEASURING STICK, HEAVY DUTY, ADULT Weight increments/graduation of 50 grams	Weight increments/graduation of 100 grams



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

**Revision and clarification to provisions/specifications in the Bidding Documents:**

Item No. 7: 2 UNITS WEIGHING SCALE WITH HEIGHT MEASURING STICK, HEAVY DUTY, ADULT

With initial calibration report or certification

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23<sup>rd</sup> day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 30<sup>th</sup> day of May 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli", is written over the "Approved by:" text.

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health

**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

Item No. <b>I</b>	<b>DENTAL INSTRUMENT CABINET</b>	Qty./Unit	<b>2 Units</b>
----------------------	----------------------------------	-----------	----------------

Name of Manufacturer:	Country of Origin
-----------------------	-------------------

Brand:	Model: (if applicable)
--------	------------------------

ABC: <b>50,000.00</b>	
-----------------------	--

PURCHASER'S SPECIFICATION	STATEMENT OF COMPLIANCE
---------------------------	-------------------------

<p><b>A. Technical Specifications:</b></p> <ul style="list-style-type: none"> <li>● Stainless steel grade 304</li> <li>● Material: Body -Stainless Steel / Table - Tempered Glass</li> <li>● No. of Drawers: At least Five (5)</li> <li>● Size: at least HxLxD 820mm x 500mm x 500mm</li> <li>● Accessories:             <ul style="list-style-type: none"> <li>- Silent anti-skid wheel with brakes</li> <li>- Anti-corrosion, fireproof, high hardness, tempered glass top with baffle.</li> <li>- 5 suitable drawer and compartment for different demand</li> <li>- Separating plate</li> </ul> </li> </ul>	
--	--

<p><b><u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u></b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30 calendar days</b> upon receipt of Notice to Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Warranty:</b> Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. With "<b>DOH-MMCHD HFEP</b>"(Government Property not for sale) sticker in each unit.</p> <p>X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X</p> <p>Ospital ng Paranaque II – 2 units</p>	
---	--



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

the warranty period.

5. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each.

X-X

Recipient:

Ospital ng Paranaque II - 7 units

**C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

Item No. <b>5</b>	<b>ER BED / TRANSFER STRETCHER</b>	Qty./Unit	<b>2 Units</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>140,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> <li>• Safe Working Load: At least 300 kg</li> <li>• Height Adjustment: Can accommodate from 580 mm to 870 mm</li> <li>• Trendelenburg/Reverse Trendelenburg: -18° to 18°</li> <li>• Protection Class: IPX4 or better</li> <li>• Width Siderails Down: 700mm to 800mm</li> <li>• Width Siderails Up: 800mm to 850mm</li> <li>• Total Length: 2100mm to 2200mm</li> <li>• Mattress Width: 650mm to 700mm</li> </ul>			



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- Mattress Thickness: At least 75mm
- Siderails Height: At least 355mm
- Siderails Length: At least 1190mm
- Features:
  - Made up of steel-reinforced frame.
  - Fully integrated side rails with gap at the head.
  - With hydraulic control pedals for height adjustment on both sides.
  
  - With control pedal for Trendelenburg, Reverse Trendelenburg and Zero leveling on both sides.
  - With push handle at the head end.
  - Fold-down side rails.
  - With four (4) corner roller bumpers.
  - With brake pedals at head and foot end.
  - Integrated IV pole.
- Standard Accessories:
  - Oxygen tank holder
  - IV pole (2 hooks or more)
  - Transfer Mattress

### **B. REQUIREMENTS IF AWARDED THE CONTRACT:**

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within **30 calendar days** upon receipt of Notice to Proceed.

2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

3. **Warranty:** Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

5. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English lang



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p>b) Operation manual in English language</p> <p>6. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>x-Recipient: Ospital ng Paranaque II - 2 units</p>	
<p><b><u>C Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p>	
<p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	

<p>Republic of the Philippines Department of Health Metro Manila Center for Health Development</p> <p><b>TECHNICAL SPECIFICATIONS</b></p>			
<p>Item No. <b>6</b></p>	<p><b>DELIVERY TABLE</b></p>	<p>Qty./Unit</p>	<p><b>8 Unit</b></p>
<p>Name of Manufacturer:</p>		<p>Country of Origin</p>	
<p>Brand:</p>		<p>Model: (if applicable)</p>	
<p>ABC: <b>480,000.00</b></p>			
<p>PURCHASER'S SPECIFICATION</p>		<p>STATEMENT OF COMPLIANCE</p>	
<p><b>A. Technical Specifications:</b></p> <ul style="list-style-type: none"> <li>• Purpose: Gynecological diagnosis, operation, parturition, clinic</li> <li>• Dimensions: At least 1900 mm x 550 mm x 800 mm</li> <li>• Backrest lifting: 0-50°</li> <li>• Footrest lifting: 0-75°</li> <li>• Three sections: Backrest and footrest are adjustable</li> <li>• Material: Stainless Steel</li> <li>• Table surface: Leather</li> <li>• Inner Material: High density sponge</li> <li>• With 2 pcs of footplate cover with leather and sponge</li> </ul>			



<ul style="list-style-type: none"> <li>• Standard Accessories               <ul style="list-style-type: none"> <li>- 1 pair of footrest</li> <li>- 1 pc mattress</li> <li>- 1 pc dirt bucket stainless steel with wheels</li> <li>- 1 pc sundries shelf</li> <li>- Stainless steel grade 304</li> <li>- Stool, Stainless Steel (Height adjustable)</li> </ul> </li> </ul>	
<p><b>B. <u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> <li>a) Service manual in English language</li> <li>b) Operation manual in English language</li> </ul> <p>6. With <b>"DOH-MMCHD HFEP"</b> (Government Property not for sale) sticker in each unit</p> <p>X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X</p> <p>Health Facilities of Paranaque - 6 units</p> <p>Ospital ng Paranaque II – 2 units</p>	
<p><b><u>C Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon</p>	



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 7	<b>WEIGHING SCALE WITH HEIGHT MEASURING STICK, HEAVY DUTY, ADULT</b>	Qty./Unit	<b>2 Units</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>30,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<b>A. Technical Specifications:</b>  Mechanical weighing scale with height rod 220 kgs maximum weighing capacity Non-slip steel platform Adjustable/telescopic height rod with 1 mm graduation Height rod can measure heights from 0.6 meter up to 2 meters high  Engraved numbers on eye level scale beam and height rod Large low-rise platform Weight increments/graduation of 100 grams Height rod of aluminum material with plastic head positioner Powder coated body "No sharp edges" construction Built in transport castors With initial calibration report or certification			
<b>B. REQUIREMENTS IF AWARDED THE CONTRACT:</b>  1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30 calendar days</b> upon receipt of Notice to Proceed.  2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical			





# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"><li>a) Service manual in English language</li><li>b) Operation manual in English language</li></ul> <p>6. With "<b>DOH-MMCHD HFEP</b>" (<b>Government Property not for sale</b>) sticker in each unit</p> <p>X-X</p> <p>Recipient: Hospital ng Paranaque I - 2 units</p>	
<p><b><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	