



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2022 – 088E**

### PROCUREMENT OF 17 UNITS CAUTERY MACHINE (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
*No changes as stipulated in technical specifications	
<b>FROM</b>	<b>TO</b>

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23<sup>rd</sup> day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 30<sup>th</sup> day of May 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

**Technical Specifications**

Item No.1	<b>CAUTERY MACHINE</b>	Qty./Unit	<b>17 Units</b>
-----------	------------------------	-----------	-----------------

Name of Manufacturer:	Country of Origin
-----------------------	-------------------

Brand:	Model: (if applicable)
--------	------------------------

ABC: <b>850,000.00</b>	
------------------------	--

PURCHASER'S SPECIFICATION	STATEMENT OF COMPLIANCE
---------------------------	-------------------------

<p><b>A. Technical Specifications:</b></p> <ul style="list-style-type: none"> <li>• Primary Power Requirement: 220V, 50/60Hz</li> <li>• Functional Characteristics:</li> <li>• Mode    Max Power (W)    Max Voltage (Vpp)    Rated Load (Ohms)    Freq. (kHz)</li> <li>HI        35                    8,000                    1,000</li> <li>24.4</li> <li>LO        20                    3,000</li> <li>500       32.3</li> <li>BI        35                    3,000</li> <li>500       32.3</li> <li>• Power Accuracy at Rated Load</li> <li>- LO and BI: 0 to 10W; Greater than 90mW or 10% of setting</li> <li>- 10 to maximum setting: 10% of setting</li> <li>- HI: Greater of 1W or 10% of setting</li> <li>• RF Frequency: 450kHz ± 50kHz (Open Circuit)</li> <li>RF Frequency: 450kHz ± 50kHz (Open Circuit)</li> <li>• Output Wave Shape: Clipped, Damped Sinusoidal</li> <li>• Line Regulation: ±5% at max. power for a ±10% line voltage</li> <li>• Duty Cycle: Intermittent: 30 sec ON, 30 sec OFF</li> <li>• Audio: Adjustable from &gt;45dBA to &gt;65dBA at 1 meter</li> <li>• Display Setting</li> <li>- Low Output Terminal</li> <li>0-10 range: 0.1-watt increments</li> <li>10-20 range: 1-watt increments</li> <li>- Bipolar &amp; High Output Terminals</li> </ul>	
--	--

