



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 089E

PROCUREMENT OF 18 SETS UPRIGHT FREEZER 14 CU. FT. (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Dual function: Freezer or Chiller	Dual function: Freezer and Chiller
Units	Sets

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23rd day of May 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 30th day of May 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development Technical Specifications			
Item No.1	UPRIGHT FREEZER 14 CU. FT.	Qty./Unit	18 Sets
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 720,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • Capacity: At least 14 cu. ft. • Dimensions: Manufacturer's Standard • Material: Stainless Steel • No frost • Dual function: Freezer and Chiller • Digital Temperature Control • Chiller: At least 2°C to 10°C • Freezer: At least -16°C to -24°C • With at least 3 shelves • With Door Alarm 			
B. <u>REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 4. Notarized undertaking that the supplier shall conduct the			

