



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 102E

PROCUREMENT OF ONE (1) UNIT COLONOSCOPE (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<ul style="list-style-type: none"> • Clamp aperture: Φ of 3.7 mm • Diameter of distal- end: Φ13.0mm • Diameter of insert tube: Φ12.8mm • Total Length: 1, 635mm • Depth of view 3-100mm • Resolution: Sony 440, 000 Pixels high resolution Photosensitive CCD •Type: Handheld -Must be compatible with the existing equipment of the hospital -Video Processor: PentaX EPK-1000 	
FROM	TO
• Insertion Tube Diameter: maximum of 11.6 mm	• Insertion Tube Diameter: maximum of 11.6 mm - removed
• Working Length: 1,700	• Working Length: 1,335mm
• Angulation: 180/180	• Angulation or Tip deflection: 180°/180° up and down
• Angulation: 160/160	• Angulation: 160°/160° right and left
• Angle of view: 140°	• Angle of view or view of field: 140°
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 working days upon receipt of Notice to Proceed.
3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.	3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 7 working days upon delivery of the equipment.
4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	4. Warranty: Warranty certificate for eighteen (18) months on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that	5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within seven (7) working days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
---	---

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 27th day of June 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of June 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli", is written over the printed name.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	COLONOSCOPE	Qty./Unit	1 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 1,560,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> • Channel Diameter: maximum of 3.8 mm • Clamp aperture: Φ of 3.7 mm • Diameter of distal- end: Φ13.0mm • Diameter of insert tube: Φ12.8mm • Working Length: 1,335mm • Total Length: 1, 635mm • Angulation or Tip deflection: 180°/180° up and down • Angulation: 160°/160° right and left • Angle of view or view of field: 140° • Depth of view 3-100mm • Resolution: Sony 440, 000 Pixels high resolution Photosensitive CCD • Product life: Minimum of 4-5 years Type: Handheld -Must be compatible with the existing equipment of the hospital -Video Processor: PentaX EPK-1000			
Requirements if awarded the Contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including			

