



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 103E

PROCUREMENT OF ONE (1) UNIT AUTOMATIC X-RAY FILM PROCESSOR (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:
--

*No changes as stipulated in technical specifications.
--

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 27th day of June 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 19th day of June 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli Annie Grace P. Sudiactal".

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	AUTOMATIC X-RAY FILM PROCESSOR (REBID)	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,800,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Processing Cycle: Adjustable 60-300 second cycle capability • Throughput: At least 240 films/hour (35x43 cm) • Dimensions: Manufacturer's Standard • Frames: can process sheet films up to 35 x 43 cm • Tank Volumes: <ul style="list-style-type: none"> - Developer: At least 12 L - Fixer: At least 12 L - Wash At least 12 L • Inclusions <ul style="list-style-type: none"> - X Ray Processor Stand - Two (2) Replenishment tanks, at least 20 L 			
Requirements if awarded the Contract: <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital 			

