

Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 110E

PROCUREMENT OF VARIOUS CENTRIFUGES (LINE BIDDING) (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

| Revision and clarification to provisions/specifications in the Bidding Documents: | |
|---|--|
| FROM | TO |
| <p>PROCUREMENT TITLE: PROCUREMENT OF VARIOUS CENTRIFUGES (LOT BIDDING) (REBID)</p> | <p>PROCUREMENT TITLE: PROCUREMENT OF VARIOUS CENTRIFUGES (LINE BIDDING) (REBID)</p> |
| <p>Item No. 1: ELECTRIC CENTRIFUGE, 12-PLACER</p> | <p>Item No. 1: ELECTRIC CENTRIFUGE, 12-PLACER</p> |
| <p>-Max. Speed (RPM): 15,000</p> | <p>Max. Speed (RPM): at least 6,000</p> |
| <p>-Max. RCF: 21,382 for angle-rotors</p> | <p>Max. RCF: at least 4,000</p> |
| <p>-Max. 6,000: At least Speed 6,000</p> | <p>Max. Speed at least 6,000</p> |
| <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> | <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.</p> |
| <p>2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> | <p>2. Prior to acceptance, the end user shall conduct a physical inspection. The instruments must have no physical damage and defect.</p> |
| <p>3. Warranty: Warranty certificate for one (1) year on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> | <p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> |
| <p>Item No. 2: ELECTRIC CENTRIFUGE, 24-PLACER</p> | <p>Item No. 2: ELECTRIC CENTRIFUGE, 24-PLACER</p> |
| <p>With metal lid casing, metal lid and stainless steel centrifugation chamber</p> | <p>With metal lid casing or plastic, metal lid and stainless steel centrifugation chamber</p> |
| <p>-Max. Speed (RPM): 15,000</p> | <p>Max. Speed (RPM): at least 6,000</p> |
| <p>-Max. RCF: 21,382 for angle-rotors</p> | <p>Max. RCF: at least 4,000</p> |
| <p>-Max. 6,000: At least Speed 6,000</p> | <p>Max. Speed: At least 6,000 RPM</p> |
| <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> | <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.</p> |
| <p>2. Prior to acceptance, the end user shall conduct</p> | <p>2. Prior to acceptance, the end user shall conduct</p> |



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

| | |
|--|---|
| a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. | a physical inspection. The instruments must have no physical damage and defect. |
| 3. Warranty: Warranty certificate for one (1) year on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. | 3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. |
| Revision and clarification to provisions/specifications in the Bidding Documents: | |
| Item No. 3: HEMATOCRIT CENTRIFUGE, 12-PLACER | |
| Inclusions: - 40 bottles of capillary tubes (100 pcs/bottle) - Sealing Clay | |
| FROM | TO |
| Metal hematocrit reader | Metal hematocrit reader or Manufacturer Standard |
| 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. | 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. |
| 3. Warranty: Warranty certificate for one (1) year on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. | 3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. |

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD


Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

| | | | |
|--|----------------------------------|-------------------------|-----------|
| Republic of the Philippines Department of Health Metro Manila Center for Health Development | | | |
| Technical Specifications | | | |
| Item No. 1 | ELECTRIC CENTRIFUGE, 12-PLACER ✓ | Qty./Unit | 2 UNITS ✓ |
| Name of Manufacturer: | | Country of Origin | |
| Brand: | | Model: (if applicable) | |
| ABC: 160,000.00 ✓ | | | |
| PURCHASER'S SPECIFICATION | | STATEMENT OF COMPLIANCE | |
| A. Technical Specifications: With metal lid casing, metal lid and stainless steel centrifugation chamber ✓ With imbalance shut off function ✓ With emergency lid release function ✓ With motor overheating protection ✓ Centrifuge ✓ -Dimensions: Manufacturer's Standard ✓ -Motor: Carbon free (Brushless) ✓ - Max. Speed (RPM): at least 6,000 ✓ - Max. RCF: at least 4,000 ✓ -Running time: 1 sec – 99 min ✓ -Power Supply: 220V, 60 Hz ✓ Rotor ✓ -Type: Angle rotor ✓ -Capacity: 12 placer ✓ -Can accommodate 2 ml, 5 ml and 10 ml test tubes ✓ - Max. Speed at least 6,000 ✓ -Max. RCF: At least 4000 ✓ | | | |
| <u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. ✓ 2. Prior to acceptance, the end user shall conduct a physical inspection. The instruments must have no physical damage and defect. ✓ | | | |



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

3. **Warranty:** Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
5. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
- a) Service manual in English language
 - b) Operation manual in English language
6. With “**DOH-MMCHD HFEP**” (Government Property not for sale) sticker in each unit

X-X

Recipient: Health Facilities in City of Manila – 2 units

C Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
 Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

| | | | |
|---|--------------------------------|-------------------------|---------|
| Republic of the Philippines Department of Health Metro Manila Center for Health Development Technical Specifications | | | |
| Item No. 2 | ELECTRIC CENTRIFUGE, 24-PLACER | Qty./Unit | 2 UNITS |
| Name of Manufacturer: | | Country of Origin | |
| Brand: | | Model: (if applicable) | |
| ABC: 280,000.00 | | | |
| PURCHASER'S SPECIFICATION | | STATEMENT OF COMPLIANCE | |
| <p>A. Technical Specifications:</p> <p>With metal lid casing or plastic, metal lid and stainless steel centrifugation chamber</p> <p>With imbalance shut off function</p> <p>With emergency lid release function</p> <p>With motor overheating protection</p> <p>Centrifuge</p> <ul style="list-style-type: none"> -Dimensions: Manufacturer's Standard -Motor: Carbon free (Brushless) - Max. Speed (RPM): at least 6,000 - Max. RCF: at least 4,000 -Running time: 1 sec – 99 min -Power Supply: 220V, 60 Hz <p>Rotor</p> <ul style="list-style-type: none"> -Type: Angle rotor -Capacity: 24 placer -Can accommodate 2 ml, 5 ml and 10 ml test tubes - Max. Speed: At least 6,000 RPM -Max. RCF: At least 4000 | | | |
| <p><u>B REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection. The</p> | | | |



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

| | |
|--|--|
| <p>instruments must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none">a) Service manual in English languageb) Operation manual in English language <p>6. With “DOH-MMCHD HFEP” (Government Property not for sale) sticker in each unit</p> <p>X-X</p> <p>Recipient: Health Facilities in City of Manila – 2 units</p> | |
| <p><u>c Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p> | |



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

| | | |
|---|---|--------------------------|
| Republic of the Philippines Department of Health Metro Manila Center for Health Development | | |
| Technical Specifications | | |
| Item No. 3 | HEMATOCRIT CENTRIFUGE, 12-PLACER | Qty./Unit 3 UNITS |
| Name of Manufacturer: | | Country of Origin |
| Brand: | | Model: (if applicable) |
| ABC: 330,000.00 | | |
| PURCHASER'S SPECIFICATION | | STATEMENT OF COMPLIANCE |
| A. Technical Specifications: Power Supply Input: 220 VAC, 60 Hz Motor Type: Brushless DC Motor No. of tubes: 12 placer (min) Speed Range 0 to 12,000 rpm Time Range: 0 to 60 minutes Metal body Adjustable minute timer On/Off indicator light Automatic voltage regular (compatible to the equipment power requirement) Metal hematocrit reader or Manufacturer Standard Inclusions: - 40 bottles of capillary tubes (100 pcs/bottle) - Sealing Clay With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit. | | |
| B REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Testing Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. | | |



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

| | |
|--|--|
| <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff. (If necessary)</p> <p>4. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none">a) Service manual in English languageb) Operation manual in English language <p>7. With “DOH-MMCHD HFEP” (Government Property not for sale) sticker in each unit</p> <p>X-X</p> <p>Recipient: Health Facilities in Taguig City – 3 units</p> | |
| <p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p> | |