

METRO MANILA CENTER FOR HEALTH DEVELOPMENT


SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 112E

PROCUREMENT OF VARIOUS HOSPITAL BED & TABLE (LINE BIDDING) (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
PROCUREMENT TITLE: PROCUREMENT OF VARIOUS HOSPITAL BED & TABLE (LOT BIDDING) (REBID)	PROCUREMENT TITLE: PROCUREMENT OF VARIOUS HOSPITAL BED & TABLE (LINE BIDDING) (REBID)
Item No. 1: BED, PLAIN WITH MATTRESS AND IV STAND • Material: Aluminum alloy or better	Item No. 1: BED, PLAIN WITH MATTRESS AND IV STAND • Material: Aluminum alloy, powder coated steel or better
• Side rails: at least ½ length	• Side rails: at least ½ length, collapsible/ sliding
• Cover: Leatherette cover with DOH letters printed in the center (front)	• Cover: Leatherette cover with DOH letters printed in the center (front), color green, 6” tall
• With patient name holder at foot and head part”	• With patient name holder at foot part
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end- users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end- users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1: BED, PLAIN WITH MATTRESS AND IV STAND	
• With caster wheels and brakes	
Item No. 2: BED MECHANICAL 2 CRANKS WITH MATTRESS • 2 cranks mechanism for back raise	Item No. 2: BED MECHANICAL 2 CRANKS WITH MATTRESS • 2 cranks mechanism for back raise and knee raise
• Knee raise	• Knee raise – removed
• Sliding full side rails	• Sliding / folding side rails
• Round tubular head board and foot board	• ABS Plastic head board and foot board
• Height from bed surface is at least 21”	• Height from bed surface is at least 18”
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end- users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end- users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.
2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	2. Prior to acceptance, the end user shall conduct a physical inspection. The instruments must have no physical damage and defect.



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

b) Operation manual in English language	b) Operation manual in English language - removed
Item No. 3: ELECTRIC DELIVERY TABLE 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	Item No. 3: ELECTRIC DELIVERY TABLE 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.
4. Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	4. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
Revision and clarification to provisions/specifications in the Bidding Documents:	
7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.	
Item No. 4: EXAMINATION TABLE Width: At least 20"	
6. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.	
Item No. 4: EXAMINATION TABLE 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	Item No. 4: EXAMINATION TABLE 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	BED, PLAIN WITH MATTRESS AND IV STAND	Qty./Unit	3 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 104,100.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • Dimensions: Manufacturer's Standard • Material: Aluminum alloy, powder coated steel or better • Side rails: at least ½ length, collapsible/sliding • Mattress: 4 in thick • Cover: Leatherette cover with DOH letters printed in the center (front), color green, 6" tall • Mounted on protective stumps – provision for IV rod (IV socket); 4 (one on each side) • With patient name holder at foot part • Knock-down construction • With caster wheels and brakes 			
B. REQUIREMENTS IF AWARDED THE CONTRACT: <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect. 3. Warranty: Warranty certificate for one (1) year on craftsmanship. Attaching also undertaking that the supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall 			




Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 2	BED MECHANICAL 2 CRANKS WITH MATTRESS	Qty./Unit	3 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 126,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
A. Technical Specifications: <ul style="list-style-type: none"> • 2 cranks mechanism for back raise and knee raise • Rectangular tubular frame • Sliding / folding side rails • IV pole & socket • ABS Plastic head board and foot board • Caster Wheel w/ 4" diameter • Foot lock on all casters (directional) • Leatherette cover mattress -at least 35" width x 4" thick x 76" length • with patient name holder at foot park • Height from bed surface is at least 18" 			
B. <u>REQUIREMENTS IF AWARDED THE CONTRACT:</u> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Prior to acceptance, the end user shall conduct a physical inspection. The instruments must have no physical damage and defect. 3. Warranty: Warranty certificate for one (1) year on craftsmanship. Attaching also undertaking that the supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 3	ELECTRIC DELIVERY TABLE	Qty./Unit	1 Set
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 390,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
A. Technical Specifications: <ul style="list-style-type: none"> - Length: 1900±50mm - Width: 600±10mm - Height: (750-950) ±20mm - Back section fold upward ≥60° - Forward: ≥10° Backward ≥20° - Back Section: Manufacturer's Standard - Seat Section: Manufacturer's Standard - Auxiliary: Manufacturer's Standard - Features - Integrated obstetric bed is suitable for conducting parturient delivery, gynecological surgery, operative abortion, diagnosis and examinations. - Various actions are controlled by remote controller - The bed surface and base both adopt high grade stainless steel - Equipped with stainless steel sewage basin and other accessories - Foot treadle brake device which is easy for moving operating table. - With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit. 			
B. REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 4	EXAMINATION TABLE	Qty./Unit	12 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 300,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
A. Technical Specifications: Steel top 3 section adjustable With stainless steel heel detachable stirrups, Stainless steel: at least grade 304 With mattress and leatherette cover at least 4-inch With stainless steel drain pan, Stainless Steel: at least grade 304 With retractable rubber covered footstep Table with at least 3 drawers and 1 cabinet, Material: Powder Coated or Better Drawer & cabinet dimensions: Manufacturer's Standard Width: At least 20"			
B. REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Testing Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the			

