



Republic of the Philippines
 Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 113E

PROCUREMENT OF VARIOUS MEDICAL EQUIPMENT (LOT BIDDING) (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1: FETAL MONITOR	
- 15 pcs. Thermal Paper	
FROM	TO
<ul style="list-style-type: none"> • Range: 30-240 bpm, 50-210 bpm 	<ul style="list-style-type: none"> • Range: 30-240 bpm, 50 - 210 bpm or 60 – 210 bpm
Item No. 1: FETAL MONITOR <ul style="list-style-type: none"> • Approx. 7-inch. TFT color LCD Display 	Item No. 1: FETAL MONITOR <ul style="list-style-type: none"> • At least 7-inch TFT color LCD Display
- 1 unit UC probe	1 unit UC (Uterine contraction) probe or equivalent
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.
Item No. 2: FETAL DOPPLER <ul style="list-style-type: none"> • Standard Accessories: - 1 conductive gel - 1 ultrasonic probe 	Item No. 2: FETAL DOPPLER <ul style="list-style-type: none"> • Standard Accessories: - 1 gallon of conductive gel - 1 ultrasonic probe with spare
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD


Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
 Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	FETAL MONITOR	Qty./Unit	1 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 160,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • Dual pulsed Doppler FHR ✓ • Ultrasound Frequency: 1 MHz ✓ • Range: 30-240 bpm, 50-210 bpm or 60 – 210 bpm ✓ • With auto-detection of dual fetal movement ✓ • With array-type thermal printer ✓ • With automatic printing function ✓ • At least. 7-inch. TFT color LCD Display ✓ • 3-channel display: FHR I, FHR II, UC • With Clinical event marking • Up to 150 patient data saving • Dimensions: Manufacturer's Standard • Can be connected to a central monitoring system through PC interface • Accessories <ul style="list-style-type: none"> - 2 units Ultrasound Doppler probe - 1 unit event marker jack - 3 units probe belt - 1 unit UC (Uterine contraction) probe or equivalent - 1 unit ultrasound gel - 2 units probe holder - Battery 			
B. REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 2	FETAL DOPPLER	Qty./Unit	10 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 450,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> ● Power Supply: Dual Power Operation (220VAC, 60 Hz and Battery should be rechargeable ● Type: Table Top (heavy duty) ● Material: Made of hard plastic/ ABS Plastic ● Standard Features: <ul style="list-style-type: none"> - LCD Display or better - Fetal Heart Rate - Built-in loudspeaker: with noise reduction - Probe holder ● Standard Accessories: <ul style="list-style-type: none"> - 1 gallon of conductive gel - 1 ultrasonic probe with spare 			
B. REQUIREMENTS IF AWARDED THE CONTRACT: <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 			

