



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 114E

PROCUREMENT OF 6 UNITS REAGENT REFRIGERATOR (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
With LED lamp	
FROM	TO
<ul style="list-style-type: none">• Stainless steel interior	<ul style="list-style-type: none">• Stainless steel interior• Interior Material: Cold rolled steel, power coated
<ul style="list-style-type: none">• Cyclical defrosting and evaporator temperature detection system	<ul style="list-style-type: none">• Cyclical defrosting and evaporator temperature detection system or auto defrost
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson




Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	REAGENT REFRIGERATOR	Qty./Unit	6 UNITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 210,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> • Power Supply Input: 220V, 50/60Hz • Effective Capacity: at least 300 Liters • Polyurethane foam insulation • Hinged door, double-paned glass door • Compressor, at least 160 W • Interior Material: Cold rolled steel, powder coated • Cyclical defrosting and evaporator temperature detection system or auto defrost • Digital display temperature • Temperature range of 2°C to 14°C • With fluorescent lamp • With 4 caster wheel swiveling with brakes • Safety Features: <ul style="list-style-type: none"> - High and Low Temperature alarm system with audio-visual indicator - Temperature recorder - Door ajar alarm • With LED lamp 			
<p>B. <u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none">a) Service manual in English languageb) Operation manual in English language <p>6. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.</p> <p>x-x</p> <p>Recipient: Health Facilities in Taguig City – 6 units</p>	
<p>C. <u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	