



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 140E

PROCUREMENT OF 1 UNIT LOW TEMPERATURE PLASMA STERILIZER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Additional • Front LED process monitoring display • Sterilization Agent: 50% Hydrogen Peroxide (7cycle/cassette) • AVR (Automatic Voltage Regulator)	
• With electric power alarm • With at least 7” Touch Screen LCD monitor, colored • Has Embedded Computer (MS Windows System)	• With electric power failure alarm • With at least 7” Touch Screen LCD monitor, colored - retained • Has Embedded Computer (MS Windows System) - retained
- 1 Cassette Sterilant Sterilizing Agent at Least 7 cycle/cassette	- 1 Cassette Sterilant Sterilizing Agent at Least 7 cycle/cassette - retained
• Inclusions: 1 Unit Documentation Label System	• Inclusions: 1 Unit Documentation Label System or its equivalent
• Inclusions: 1 Roll Plasma Adhesive Tape for Plasma with documentation label	• Inclusions: 1 Roll Plasma Adhesive Tape for Plasma with documentation label or its equivalent
• Inclusions: - 1 roll per Size with Sealer, 75mm x 100m, 100mm x 100m, 150mm x 100m, 200mm x 100m, 250mm x 100m, 300mm x 100mm)	• Inclusions: - 1 roll per Size with Sealer, 75mm x 100m, 100mm x 100m, 150mm x 100m, 200mm x 100m, 250mm x 100m, 300mm x 100mm) - retained
REQUIREMENTS IF AWARDED THE CONTRACT: 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	REQUIREMENTS IF AWARDED THE CONTRACT: 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning with Preventive Maintenance and Calibration within the warranty period.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 1st day of December 2022, 9:00 AM:



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 23rd day of October 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli Annie Grace P. Sudiacal".

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	LOW TEMPERATURE PLASMA STERILIZER	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,400,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Volume: 50 liters, 2 tiered shelves, Rectangular Type • With electric power failure alarm • With at least 25 minutes cycle time • Parameters: Temperature, Pressure, Cycle time, Daily Cycle • With at least 7" Touch Screen LCD monitor, colored • Front LED process monitoring display • With USB data backup and Ethernet connection • Power Supply: 220V/60Hz • Material: Aluminum or Stainless Steel Chamber • Has Self Diagnosis Function • Has Embedded Computer (MS Windows System) • Sterilization Agent: 50% Hydrogen Peroxide (7cycle/cassette) • Inclusions: <ul style="list-style-type: none"> - 1 Cassette Sterilant Sterilizing Agent at Least 7 cycle/cassette - 1 box 250 pcs. Chemical indicator Strip for Plasma - 2 box 50pcs/box vial Biological indicator for Plasma with incubator - 1 roll per Size with Sealer, 75mm x 100m, 100mm x 100m, 150mm x 100m, 200mm x 100m, 250mm x 100m, 300mm x 100mm) 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<ul style="list-style-type: none"> - 1 Roll Therma Printer Paper - 1 Unit instrument Tray - 1 Unit Documentation Label System or its equivalent - 1 Roll Plasma Adhesive Tape for Plasma with documentation label or its equivalent <p>Device can be sterilized with Sterilization Programs: Short, Standard, Advanced including:</p> <ul style="list-style-type: none"> - Non-lumen Endoscope - Defibrillator - Ocular Lens - Camera - GI Endoscope - Cytoscope - Ureteroscope - Hysteroscope - Bronchoscope <p>Additional: AVR (Automatic Voltage Regulator)</p>	
<p>REQUIREMENTS IF AWARDED THE CONTRACT:</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60-90 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment: 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning with Preventive Maintenance and Calibration within the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period. 6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following: <ul style="list-style-type: none"> a) Service manual in English language 	

