ORDAN SHEPHILIPA IN THE SHEET SHEET

Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 141E PROCUREMENT OF 3 UNITS DENTAL UNIT, MOTORIZED WITH ACCESSORIES

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

TO
- Three (3) pcs hand piece attachments (2 hispeed adaptors & 1 low-speed adaptor)
- Auto-cup-filler & bowl flush (water flushing control valve, cup-filler with control valve)
REQUIREMENTS IF AWARDED THE CONTRACT
1. Completion Period: The delivery, installation testing and commissioning of the equipment and its accessories, including the training of endusers and maintenance staff must be completed within 30 to 60 calendar days upon receipt of
I

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 1st day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 23rd day of October 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH

Director III / BAC Chairperson



Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS		
tem No. 1 DENTAL UNIT, MOTORIZED WITH ACCESSORIES	Qty./Unit	3 Units
Name of Manufacturer:		Country of Origin
Brand:		Model: (if applicable)
ABC: 540,000.00		
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE
DENTAL UNIT, MOTORIZED WITH ACCESSORIES		
- Fully Motorized with pre-set functions		
- Touch control chair		
- Slim ergonomic design upholstery		
- LED light with sensor switch		. 19
- Double side armrest		
- Colored glass/ ceramic bowl, rotating detachable cuspidor strainer)	bowl (with plastic	
- Two (2) pcs. 3 way syringe		•
- Three (3) pcs hand piece attachments (2 hi-speed adaptor: adaptor)	s & 1 low-speed	
- Auto-cup-filler & bowl flush (water flushing control valve, control valve)	cup-filler with	
- Saliva ejector-built-in motor pump type, does not use air o	r water	
- Clean water system		
- LED X-ray film viewer (Peri-apical X-ray viewer)		
- Multi-functional foot control		
- Water tank for distilled water for hand piece		
Two (2) hi-speed hand piece		
One (1) low-speed complete with straight hand piece & Co	ntra-angle	
LED light cure unit (external)		
Tissue box with tray and cup dispenser		

STATE MENT WHEN

Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- Silent type and oil-free dental air compressor
- Saddle-type with backrest dentist stool
- With acrylic barrier

Requirements if awarded the Contract

- 1. **Completion Period**: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within **30 to 60 calendar days** upon receipt of Notice to Proceed.
- 2. **Testing**: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
- 3. **Training**: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.
- 4. **Warranty**: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
- 5. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- 6. **Manuals**: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
- 7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

Source of Fund: SAA 2021-02-0209 (HFEP 2021 CONAP)

Recipient: Health Facilities in Taguig City

Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.