



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/BID BULLETIN NO. 1

IB 2022 – 143E

PROCUREMENT OF 37 UNITS STETHOSCOPE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Chestpiece Finish: Brass/Machined Stainless Steel/Matte/Polished Stainless Steel	Chestpiece Finish: Brass or Machined Stainless Steel or Matte or Polished Stainless Steel
Chestpiece Technology: Single-sided	Chestpiece Technology: Single-sided - retained
Diaphragm Material: Epoxy/Fiberglass	Diaphragm Material: Epoxy/Fiberglass or equivalent
Diaphragm Type: Tunable Diaphragm	Diaphragm Type: Tunable Diaphragm – retained
Headset Material: Anodized Aluminum	Headset Material: Anodized Aluminum or its equivalent
Includes Special Procedures Adaptor for Pediatric and Neonatal auscultation	Includes Special Procedures Adaptor for Pediatric and Neonatal auscultation or its equivalent

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 1st day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 23rd day of October 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	STETHOSCOPE	Qty./Unit	37 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 925,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> ● Tubing should not be made of natural rubber latex and does not contain phthalate plasticizers. ● Binaural Construction: Double Lumen ● Chestpiece Finish: Brass or Machined Stainless Steel or Matte or Polished Stainless Steel ● Chestpiece Technology: Single-sided ● Diaphragm Material: Epoxy/Fiberglass or equivalent ● Diaphragm Type: Tunable Diaphragm ● Eartip Type: Soft Sealing ● Headset Material: Anodized Aluminum or its equivalent ● Includes Special Procedures Adaptor for Pediatric and Neonatal auscultation or its equivalent 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 4. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each 			

