



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 144E

PROCUREMENT OF 1 UNIT AUTOMATIC X RAY FILM PROCESSOR

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
• Processing Cycle: Adjustable 60-300 second cycle capability	• Processing Cycle: Adjustable 90, 105, 135 second cycle capability
• Throughput: At least 240 films/hour (35x43 cm)	• Throughput: For 90 seconds cycle 72 films/hour (35x43 cm) For 105 seconds cycle 62 films/hour For 135 seconds cycle 48 films/hour
• Frames: can process sheet films up to 35 x 43 cm	• Frames: can process sheet films up to 35 x 43 cm - retained

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 1st day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 23rd day of October 2022 in MMCHD

Approved by:

ALELI ANNE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	AUTOMATIC X RAY FILM PROCESSOR	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,800,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Processing Cycle: Adjustable 90, 105, 135 second cycle capability • Throughput: For 90 seconds cycle 72 films/hour (35x43 cm) For 105 seconds cycle 62 films/hour For 135 seconds cycle 48 films/hour • Dimensions: Manufacturer's Standard • Frames: can process sheet films up to 35 x 43 cm • Tank Volumes: <ul style="list-style-type: none"> - Developer: At least 12 L - Fixer: At least 12 L - Wash At least 12 L • Inclusions <ul style="list-style-type: none"> - X Ray Processor Stand - Two (2) Replenishment tanks, at least 20 L 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance 			

