



# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

## SUPPLEMENTAL/BID BULLETIN NO. 1

IB 2022 – 145E

### PROCUREMENT OF 1 UNIT SLIT LAMP

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Additional: AVR (Automatic Voltage Regulator)	
<b>FROM</b>	<b>TO</b>
• Filter: Blue, Green, (red-free), Grey, Diffusor, Yellow	• Filter: Blue, Green, (red-free), Grey, Diffusor, Yellow (UV Cut, IR Cut, Amber Cut) or its equivalent
• Magnification - 5/8/12/20/32x with 10x eyepiece - 6/10/16/25/40x with 12.5x eyepiece	Magnification - 5/8/12/20/32x with 10x eyepiece or - 6/10/16/25/40x with 12.5x eyepiece
REQUIREMENTS IF AWARDED THE CONTRACT 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS IF AWARDED THE CONTRACT 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 1<sup>st</sup> day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 23<sup>rd</sup> day of October 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. <b>1</b>	<b>SLIT LAMP</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>650,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> <li>• Designed for:             <ul style="list-style-type: none"> <li>- Routine Observation</li> <li>- Diagnosis</li> <li>- Documentation of the anterior eye segment</li> </ul> </li> <li>• Light source below stereomicroscope</li> <li>• Controls around slit projector</li> <li>• Design: Integrated</li> <li>• Slit Illumination: Halogen</li> <li>• Filter: Blue, Green, (red-free), Grey, Diffusor, Yellow (UV Cut, IR Cut, Amber Cut) or its equivalent</li> <li>• Magnification             <ul style="list-style-type: none"> <li>- 5/8/12/20/32x with 10x eyepiece or</li> <li>- 6/10/16/25/40x with 12.5x eyepiece</li> </ul> </li> <li>• Power supply: 220V, 50/60Hz</li> <li>• AVR (Automatic Voltage Regulator)</li> </ul>			
<b>Requirements if awarded the Contract</b> 1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30 to 60 calendar</b> days upon receipt of Notice to Proceed. 2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no			



# METRO MANILA CENTER FOR HEALTH DEVELOPMENT