



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 148E

PROCUREMENT OF 1 UNIT COLONOSCOPE (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Additional: Total length: 1,635mm Material – fiber Type – flexible Camera – 440, 000 Pixels high resolution Photosensitive CCD Clamp aperture: Φ of 3.7 mm Diameter of distal- end: Φ 13.0mm Depth of view 3-100mm -Must be compatible with the existing equipment of the hospital (PentaX EPK-1000)	
FROM	TO
• Insertion Tube Diameter: maximum of 11.6 mm	Insertion Tube Diameter: 12.8mm
• Working Length: 1,700	Working Length: 1,335mm
• Angulation: 180/180	Angulation or Tip deflection: 180°/180° up and down
• Angulation: 160/160	Angulation: 160°/160° right and left
• Angle of view: 140°	Angle of view or view of field: 140°
• Product life: Minimum of 4-5 years	• Product life: Minimum of 4-5 years - removed
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 1st day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

For guidance and information of all concerned.

Issued this 23rd day of October 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli", is written over the printed name.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	COLONOSCOPE	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,560,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Channel Diameter: maximum of 3.8 mm • Insertion Tube Diameter: 12.8 mm • Working Length: 1,335mm • Total Length: 1,635mm • Angulation or Tip deflection: 180°/180° up and down • Angulation: 160°/160° right and left • Angle of view or view of field: 140° • Material – fiber • Type – flexible • Camera – 440, 000 Pixels high resolution Photosensitive CCD • Clamp aperture: Φ of 3.7 mm • Diameter of distal- end: Φ13.0mm • Depth of view 3-100mm -Must be compatible with the existing equipment of the hospital (PentaX EPK-1000) 			
REQUIREMENTS IF AWARDED THE CONTRACT: <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance 			

