



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 160E**

**PROCUREMENT OF 1 UNIT HEMATOLOGY ANALYZER**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>
--

*No changes as stipulated in technical specifications
---

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13<sup>th</sup> day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of December 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli", is written over the printed name.

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>HEMATOLOGY ANALYZER</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>1,000,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>TECHNICAL SPECIFICATIONS:</b> <ul style="list-style-type: none"> <li>• System: Fully Automated 5-part differential</li> <li>• Measurement Principle: Fluorescence staining and Flow cytometry</li> <li>• Measurement Parameters: 28 parameters, including 4 research parameters</li> <li>• Throughput: At least 60 tests/hour</li> <li>• Test Mode: CBC/CBC-Diff</li> <li>• Sample Type: Whole blood/Capillary Blood/Pre-dilution blood</li> <li>• Sampling Method: Manual Sampling</li> <li>• Sample Volume: 20ul</li> <li>• Sample pre-dilution: Manual injection/Automatic injection</li> <li>• Reagents: 4 routine reagents (Diluent, Lysing reagent, Hemolysing reagent, Fluorescent Dye) 1 cleaning solution</li> <li>• Quality Management               <ul style="list-style-type: none"> <li>- Calibrator should be FDA registered to ensure safety and quality of the machine.</li> </ul> </li> <li>• Data Storage: At least 200,000</li> <li>• Inclusion:               <ul style="list-style-type: none"> <li>- One (1) set of start-up reagents and accessories with GHL (Glucose, Hemoglobin and Lactate) meter kit</li> <li>- One (1) computer set</li> </ul> </li> </ul>			

