



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 161E**

**PROCUREMENT OF 1 SETS LAPAROSCOPIC MACHINE INSTRUMENTS WITH  
MEDICAL RECORDER**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
The following instruments must be compatible with the existing tower:	The following instruments must be compatible with OUMAN OM-910HDA
<ul style="list-style-type: none"> <li>• 20 cannister Wipes for instruments effective in 1 minute contact time. Belzankonium chloride 0.310%, didecyldimethylammoniumchloride0.940%, polyhexamethylenebiguanide hydrochloride 0.866%</li> </ul>	<ul style="list-style-type: none"> <li>• 20 cannister Wipes for instruments effective in 1 minute contact time. Benzalkonium chloride 0.310%, didecyldimethylammoniumchloride0.940%, polyhexamethylenebiguanide hydrochloride 0.866%</li> </ul>
<b>REQUIREMENTS IF AWARDED THE CONTRACT:</b> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	<b>REQUIREMENTS IF AWARDED THE CONTRACT:</b> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30-60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13<sup>th</sup> day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of December 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<b>Republic of the Philippines</b> <b>Department of Health</b> <b>Metro Manila Center for Health Development</b>			
<b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	LAPAROSCOPIC MACHINE INSTRUMENT WITH MEDICAL RECORDER	Qty./Unit	1 SET
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 1,500,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <p>The following instruments must be compatible with OUMAN OM-910HDA</p> <p>Laparoscopic Instruments</p> <ul style="list-style-type: none"> <li>● 1 pc Laparoscopy Scope 10mm dia x 30 degree. 330mm</li> <li>● 2 pcs Trocar sleeve with stop-cock &amp; obturator pyramidal 6mm</li> <li>● 1 pc Trocar sleeve with stop-cock &amp; obturator Pyramidal 11mm</li> <li>● 2 pcs Hasson Trocar complete 11mm</li> <li>● 1 pc Reducer removable with holder, from 10mm – 5mm</li> <li>● 1 pc Veress needle 2mm length 200mm</li> <li>● 1 pc Maryland dissecting forceps 5mm</li> <li>● 1 pc Blunt dissecting and grasping forceps 5mm</li> <li>● 1 pc Mixer dissecting forceps 5mm</li> <li>● 1 pc Clip Applying forceps 10mm</li> <li>● 1 pc Bobcock grasping forceps jaw 5mm</li> <li>● 1 pc Metzenbaum scissor slightly curved</li> <li>● 1 pc Endo needle holder 5mm</li> <li>● 1 pc Knot guide with eye</li> <li>● 1 pc Spatula 5mm</li> </ul>			



- 1 pc L hook 5mm
- 1 pc Bowell grasper 5.5mm
- 1 pc Suction irrigator with sliding valve and standard tube
- 1 box (20 per box) Titanium clips
- 20 cannister Wipes for instruments effective in 1 minute contact time. Benzalkonium chloride 0.310%, didecyldimethylammoniumchloride 0.940%, polyhexamethylenebiguanide hydrochloride 0.866%

Medical Recorder

- MPEG-4AVC/H.264 G-4 AVC/H.264
- HIGH QUALITY FULL HD record (MPEG-4 AVC/H.264 zip)
- Long recording time
- Multiport (HDMI/DV, CVBS, YPbPr)
- Simplified keys and easy operation
- Still and moving picture capture
- Interface connection
  - Interface: USB2.0
  - Video input: Digital- HD-SDI and Analog- CVBS (PAL)
  - Video output HDMI/DVI/VGA
- Hardware Specification
  - Encoder: H.264 Hardware encoding
  - User interface: Recorder button. Capture button. Video source switching button. Light indicator
  - Storage Media: U Disk, 2T Max Capacity Mobile HDD

**REQUIREMENTS IF AWARDED THE CONTRACT:**

1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within **30-60** calendar days upon receipt of Notice to Proceed.
2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty

