



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 166E**

**PROCUREMENT OF 6 UNITS FETAL MONITOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
• Fetal Heart Rate Measurement Range: 30-240 bpm	• Fetal Heart Rate Measurement Range: 30-240 bpm - retained
• At least 7-inch. TFT color LCD Display	• At least 7-inch. TFT color LCD Display - retained
• At least 160 patient data saving	• At least 160 patient data saving - retained
• Power aid module: 10-12 hours	• Power aid module: 10-12 hours - retained
<b>REQUIREMENTS IF AWARDED THE CONTRACT:</b> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	<b>REQUIREMENTS IF AWARDED THE CONTRACT:</b> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13<sup>th</sup> day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of December 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development**

**TECHNICAL SPECIFICATIONS**

<b>Item No. 1</b>	<b>6 UNITS FETAL MONITOR</b>	<b>Qty./Unit</b>	<b>6 Units</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>960,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>		<b>STATEMENT OF COMPLIANCE</b>	
<p><b>TECHNICAL SPECIFICATIONS:</b></p> <ul style="list-style-type: none"> <li>• Dual pulsed Doppler FHR</li> <li>• Ultrasound Frequency: 1 MHz</li> <li>• Fetal Heart Rate Measurement Range: 30-240 bpm</li> <li>• Fetal Heart Rate Waveform Display Range: 50-210 bpm</li> <li>• With auto-detection of dual fetal movement</li> <li>• With array-type thermal printer</li> <li>• With automatic printing function</li> <li>• At least 7-inch. TFT color LCD Display</li> <li>• 3-channel display: FHR I, FHR II, UC</li> <li>• With Clinical event marking</li> <li>• At least 160 patient data saving</li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Can be connected to a central monitoring system through PC interface</li> <li>• Has built-in battery: 4-8 hours</li> <li>• Power aid module: 10-12 hours</li> <li>• Accessories               <ul style="list-style-type: none"> <li>- 2 units Ultrasound Doppler probe</li> <li>- 1 unit event marker jack</li> <li>- 3 units probe belt</li> </ul> </li> </ul>			

