



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 168E

PROCUREMENT OF 27 UNITS BP ANEROID WITH WHEELS

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:
--

*No changes as stipulated in technical specifications.
--

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13th day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6th day of December 2022 in MMCHD

Approved by:

A handwritten signature in blue ink, appearing to read "Aleli", is written over the printed name.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	BP ANEROID WITH WHEELS	Qty./Unit	27 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 599,400.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <p>Measuring range: 0 to 300mmHg</p> <p>Accuracy: ± 3mmHg</p> <p>Sub-division: 2mmHg</p> <p>Has inflatable cuff and inflation bulb</p> <p>Has bladder hose/tube and release valve</p> <p>Has ABS plastic gauge</p> <p>Gauge diameter: atleast 150mm</p> <p>Adjustable Column height: Manufacturer's standard</p> <p>Has a plastic base plate with 5 castors</p>			
<p>REQUIREMENTS IF AWARDED THE CONTRACT:</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after</p>			



testing and commissioning.

4. Notarized undertaking that the supplier shall provide the necessary replacement within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

6. 'With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit

X-X

Source of Fund: SAA 2022-02-0531