



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 40E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in the City of Malabon**

This Supplemental/Bid Bulletin No. I is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>2 UNITS AUTOCLAVE SPECIFICATION</b>  *No changes as stated in Technical Specifications	
<b>Item 2A</b> <b>2 UNIT BEDSIDE TABLE</b>  *Stainless Materials  *Must have a spill collector edge for thermofoil top or laminate top	
<b>From</b>	<b>To</b>
ITEM 2: 2 UNIT BEDSIDE TABLE	Item 2A: 2 UNIT BEDSIDE TABLE
<b>Item 2B</b> <b>2 SETS CATARACT INSTRUMENT SET</b>	
<b>From</b>	<b>To</b>
Item 2: 2 SETS CATARACT INSTRUMENT SET	Item 2B: 2 SETS CATARACT INSTRUMENT SET
<b>Item 3</b> <b>4 Units DIFFERENTIAL COUNTER</b>  *No changes as stated in Technical Specifications	
<b>Item 5</b> <b>2 Units EMERGENCY CART</b>  * No changes as stated in Technical Specifications	
<b>ITEM 6</b> <b>2 UNITS EXTRACTION CHAIR HEAVY DUTY STAINLESS</b>  *Additional Specs: Left and right arm rest	
<b>ITEM 7</b> <b>2 Units HEMATOCRIT CENTRIFUGE 12 PLACER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 8</b> <b>2 UNITS INSTRUMENT CABINET DOUBLE DOOR</b>  * No changes as stated in Technical Specifications	
<b>ITEM 9</b> <b>2 UNITS MEDICINE CABINET</b>  * No changes as stated in Technical Specifications	

**ITEM 10**

**4 UNITS ULTRAVIOLET LIGHT 4 BULB**

\*No changes as stated in Technical Specifications

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson 



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 41E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Sta Ana Hospital**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b>	
<b>1 UNIT CARDIAC MONITOR WITH CAPNOGRAPH</b>	
From	To
Resolution: 1024 x 768 pixels (XGA)	Resolution: 1024 x 768 pixels (XGA) or higher
Manual configuration with up to 8 user configurable profiles for care-specific configurations, and up to 6 user configurable display pages for each profile	Manufacturer's standard
Pivoting module frame *Features: Can be pivoted to a locked closed position, locked to an open 45° angle, or locked to a 90° open position. Support for Patient Data or its equivalent *Options: Two E-module slots and optional recorder	Pivoting module frame: Manufacturer's standard
Snapshot *5 s Waveform: 400 snapshots *ST: 10 snapshots *Events: 999 events	Snapshot: Optional
Paper Recorder *Number of recorder waveforms 4 *Horizontal resolution: 24 dots/mm (600 dpi) @ 25mm/sec * Vertical resolution 8 dots/mm (200 dpi)	Paper Recorder *Number of recorder waveforms minimum of 3 *Horizontal resolution: Manufacturer's standard * Vertical resolution: Manufacturer's standard
<b>Item 2</b>	
<b>1 Unit EMERGENCY CART</b>	
*No change as stated in Technical Specifications	
<b>Item 3</b>	
<b>2 Units ECG MACHINE</b>	
From	To
Arrhythmia Monitoring and Summary up to 15 minutes	Arrhythmia Monitoring and Summary up to at least 10 - 15 minutes
Inclusions: * Paper Spool	Inclusions: *Paper Spool (Optional)
<b>Item 4</b>	
<b>1 UNIT FETAL MONITOR</b>	
From	To
At least 7 inches wide TFT color LCD display	At least 7 inches wide TFT color LCD or LED display

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:



**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson







Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 42E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in City of Navotas**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>2 UNIT AUTOCLAVE 20L</b>	
<b>From</b>	<b>To</b>
Power Supply: 220 V - 3 Phase, 60 Hz	Power Supply: 220 V - 3 Phase, 60 Hz or Single Phase
<b>Item 2</b> <b>2 Units CENTRIFUGE</b>  * Capacity: Swing out rotor	
<b>Item 3</b> <b>1 Unit DENTAL INSTRUMENT CABINET</b>	
<b>From</b>	<b>To</b>
Color: 4 color	<b>Removed:</b> Color: 4 color
<b>Item 4</b> <b>1 UNIT DENTAL INSTRUMENT SET</b>	
<b>From</b>	<b>To</b>
Perisoteal Elevator Stainless Steel	Periosteal Elevator Stainless Steel
<ul style="list-style-type: none"><li>• Apical Elevator Right, Non-magnetic Hollow handle</li><li>• Apical Elevator Left, Non-magnetic Hollow handle</li><li>• Cryer (east), Non-magnetic Hollow handle</li><li>• Cryer (west), Non-magnetic Hollow handle</li><li>• Pedo forcep 150SK Stainless Steel</li><li>• Pedo forcep 151SK Stainless Steel</li></ul>	<b>REMOVE:</b> <ul style="list-style-type: none"><li>• Apical Elevator Right, Non-magnetic Hollow handle</li><li>• Apical Elevator Left, Non-magnetic Hollow handle</li><li>• Cryer (east), Non-magnetic Hollow handle</li><li>• Cryer (west), Non-magnetic Hollow handle</li><li>• Pedo forcep 150SK Stainless Steel</li><li>• Pedo forcep 151SK Stainless Steel</li></ul>
<b>Item 5</b> <b>1 Unit DENTAL ULTRASONIC SCALER</b>  *No changes as stated in Technical Specifications	
<b>Item 6</b> <b>1 Unit DENTAL UNIT MOTORIZED WITH ACCESSORIES</b>  *No changes as stated in Technical Specifications	
<b>Item 7</b> <b>1 Unit DIFFERENTIAL COUNTER</b>  *No changes as stated in Technical Specifications	
<b>Item 8</b> <b>1 Unit ECG MACHINE</b>  *No changes as stated in Technical Specifications	
<b>Item 9</b> <b>2 SETS EENT DIAGNOSTIC SET</b>	
<b>From</b>	<b>To</b>
Spare light bulb (LED/Xenon)	Manufacturer's standard

<b>Item 10</b> <b>4 UNIT EXAMINING TABLE</b>  Additional Technical Specification: *Hamilton Type	
<b>Item 11</b> <b>4 UNITS FETAL DOPPLER</b>  Standard Features: LCD Display or LED Display	
<b>Item 12</b> <b>2 UNITS INFANT WEIGHING SCALE, DIGITAL</b>  *No changes as stated in Technical Specifications	
<b>Item 13</b> <b>2 UNITS INSTRUMENT CABINET</b>  *No changes as stated in Technical Specifications	
<b>Item 14</b> <b>2 UNITS INSTRUMENT CABINET</b>	
<b>From</b>	<b>To</b>
Material: Stainless steel or equivalent	Material: Stainless steel
<b>Item 15</b> <b>8 UNITS MEDICINE CABINET</b>  *No changes as stated in Technical Specifications	
<b>Item 16</b> <b>4 UNITS PORTABLE UVC LIGHT STERILIZER</b>  *No changes as stated in Technical Specifications	
<b>Item 17</b> <b>2 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE</b>  *No changes as stated in Technical Specifications	
<b>Item 18</b> <b>2 UNITS REAGENT REFRIGERATOR</b>  *No changes as stated in Technical Specifications	
<b>Item 19</b> <b>2 UNITS STERILIZING UNIT</b>  *No changes as stated in Technical Specifications	
<b>Item 20</b> <b>2 Units URINE CENTRIFUGE</b>  *No changes as stated in Technical Specifications	
<b>Item 21</b> <b>1 UNIT VACCINE REFRIGERATOR</b>  *No changes as stated in Technical Specifications	
<b>Item 22</b> <b>2 UNITS WEIGHING SCALE WITH HEIGHT MEASURING STICK</b>  *No changes as stated in Technical Specifications	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson 



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 43E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in City of Las Pinas**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>2 UNIT AUTOCLAVE 20L</b>	
<b>From</b>	<b>To</b>
Power Supply: 220 V - 3 Phase, 60 Hz	Power Supply: 220 V - Single Phase
Chamber Capacity: 20 Liters	Chamber Capacity: 20 - 23 Liters
<b>Item 2</b> <b>1 Unit BIOSAFETY CABINET</b>  *No changes as stated in Technical Specifications	
<b>Item 3</b> <b>1 Unit CENTRIFUGE</b>  *No changes as stated in Technical Specifications	
<b>Item 4</b> <b>2 UNITS DENTAL UNIT</b>  *No changes as stated in Technical Specifications	
<b>Item 5</b> <b>1 Unit DIFFERENTIAL COUNTER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 6</b> <b>3 Units FETAL DOPPLER</b>	
<b>From</b>	<b>To</b>
Standard Features: LCD Display	Standard Features: LCD or LED Display
<b>ITEM 7</b> <b>5 Units INFANT WEIGHING SCALE, DIGITAL</b>  * No changes as stated in Technical Specifications	
<b>ITEM 8</b> <b>1 Unit INSTRUMENT CABINET DOUBLE DOOR</b>  * No changes as stated in Technical Specifications	
<b>ITEM 9</b> <b>2 UNIT MEDICINE CABINET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 10</b> <b>7 UNITS MINOR SURGICAL SET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 11</b> <b>7 UNITS REAGENT REFRIGERATOR</b>  * No changes as stated in Technical Specifications	



**ITEM 12****5 UNITS WEIGHING SCALE WITH HEIGHT MEASURING STICK**

\* No changes as stated in Technical Specifications

**ITEM 13****1 UNIT VACCINE REFRIGERATOR**

\* No changes as stated in Technical Specifications

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:



**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 44E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in City of Muntinlupa**

This Supplemental/Bid Bulletin No. I is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>2 UNIT AUTOCLAVE 20L</b>	
<b>From</b>	<b>To</b>
Power Supply: 220 V - 3 Phase, 60 Hz	Power Supply: 220 V - Single Phase
Chamber Capacity: 20 Liters	Chamber Capacity: 20 - 23 Liters
<b>Item 2</b> <b>1 Unit BIOSAFETY CABINET</b>	
<b>From</b>	<b>To</b>
External Dimension: 1354 x 759 x 1260 mm	External dimension: Manufacturer's Standard
Chamber dimension: 1259 x 605 x 587 mm	Chamber dimension: Manufacturer's Standard
<b>Item 3</b> <b>1 Unit DENTAL ULTRASONIC SCALER</b>  *No changes as stated in Technical Specifications	
<b>Item 4</b> <b>7 UNITS DRESSING SET</b>  *No changes as stated in Technical Specifications	
<b>Item 5</b> <b>3 Units DENTAL AEROSOL SUCTION MACHINE</b>  * No changes as stated in Technical Specifications	
<b>ITEM 6</b> <b>1 Unit ECG MACHINE</b>  * No changes as stated in Technical Specifications	
<b>ITEM 7</b> <b>3 Units EENT DIAGNOSTIC SET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 8</b> <b>8 Unit EXAMINING TABLE</b>  Additional Technical Specifications: Hamilton Type	
<b>ITEM 9</b> <b>10 UNIT FETAL DOPPLER</b>	
<b>From</b>	<b>To</b>
Standard Features: LCD Display	Standard Features: LCD or LED Display
<b>ITEM 10</b> <b>1 UNIT FETAL MONITOR</b>  * No changes as stated in Technical Specifications	

<b>ITEM 11</b> <b>11 UNITS INSTRUMENT CABINET</b>	
<b>From</b>	<b>To</b>
Dimension: 36"w x 23 1/4" diameter x 72" height or equivalent"	Dimension: Manufacturer's standard
Stainless Steel Sheet with two side glass, two glass doors with lock or equivalent"	Stainless Steel Sheet with two side glass or stainless, two glass doors with lock or equivalent"
<b>ITEM 12</b> <b>8 UNITS MEDICINE CABINET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 13</b> <b>5 SETS MINOR SURGICAL SET/CUT DOWN SET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 14</b> <b>1 UNIT PULSE OXIMETER WITH ADULT AND PEDIA PROBE</b>	
<b>From</b>	<b>To</b>
Dimension: Minimum requirements of 267mm x 186mm x 87mm (W x D x H)"	Dimension: Manufacturer's standard
<b>ITEM 15</b> <b>1 UNIT SUCTION MACHINE HEAVY DUTY</b>	
<b>From</b>	<b>To</b>
Noise: not more than 55 Db	Noise: not more than 60 Db
Pumping rate: 20 L/min.	Pumping rate: 20 - 35 L/min.
Reservoir Material: Glass	Reservoir Material: Glass or Polycarbonate
<b>ITEM 16</b> <b>1 UNIT DIFFERENTIAL COUNTER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 17</b> <b>2 UNITS ER BED/TRANSFER STRETCHER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 18</b> <b>1 UNIT DISSOLVED OXYGEN METER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 19</b> <b>1 UNIT BENCHTOP ION SELECTIVE METER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 20</b> <b>5 UNITS ION ELECTRODES</b> <b>(Nitrates, Flourides, Chlorides and Sulfates)</b>  * No changes as stated in Technical Specifications	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:

  
**ALELI ANNE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson 





## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 45E

#### Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in the City of Pasay

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:

**Item 1**

**5 UNIT MICROSCOPE BINOCULAR**

**\*No changes as stated in Technical Specifications**

**Item 2**

**1 Unit VACCINE REFRIGERATOR**

Additional Requirements:

1.) -70°C to -80°C / 490L Capacity Upright 17.3 Cu FT.

Or

2.) 2 in 1 Freezer

Upper Portion – Ultra Low ( -70°C to -80°C)

Lower Portion – Moderate Low (-20°C to -40°C)

3.) -20°C to -40°C / 490L Capacity Upright 17.3 Cu FT.

Inclusion:

1.) Automatic Voltage Regulator

2.) Indoor Thermometer w/ Hygrometer

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 46E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in the Municipality of Pateros**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>3 UNITS FETAL DOPPLER</b>	
<b>From</b>	<b>To</b>
Large backlight LCD display	Large backlight LED display
<b>Item 2</b> <b>2 Units Examining Table</b>  *No changes as stated in Technical Specifications	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson 



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 47E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in the Quezon City**

This Supplemental/Bid Bulletin No. I is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>17 UNITS FETAL DOPPLER</b>	
<b>From</b>	<b>To</b>
Power Supply: 220 VAC, 60Hz	Power Supply: Battery Operated
<b>Item 2</b> <b>7 UNITS AUTOCLAVE 20L</b>	
<b>From</b>	<b>To</b>
Power Supply: 220 V - 3 Phase, 60 Hz	Power Supply: 220 V - Single Phase,
<b>Item 3</b> <b>5 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE</b> * No changes as stated in Technical Specifications	
<b>Item 4</b> <b>8 Units EENT DIAGNOSTIC SET</b> *No changes as stated in Technical Specifications	
<b>Item 5</b> <b>6 UNITS INFANT WEIGHING SCALE</b> * No changes as stated in Technical Specifications	
<b>ITEM 6</b> <b>1 Unit DENTAL ULTRASONIC SCALER</b> * No changes as stated in Technical Specifications	
<b>ITEM 7</b> <b>6 Unit WEIGHING SCALE WITH HEIGHT MEASURING STICK ADULT</b> * No changes as stated in Technical Specifications	
<b>ITEM 8</b> <b>1 Unit DELIVERY SET</b>	
<b>From</b>	<b>To</b>
1 pc. Sterilizing tray (Perforated)	To fit all instruments in Sterilizing tray (Perforated)
<b>ITEM 9</b> <b>5 Unit DRESSING SET</b> Additional Accessories:  *Stainless steel tray *Dimension: Manufacturer's standard or to fit all the instruments inside the tray	
<b>ITEM 10</b> <b>3 Unit CERVICAL INSPECTION SET</b>  Additional Accessories:	



*Stainless steel tray	
*Dimension: Manufacturer's standard or To fit all the instruments inside the tray	
<b>ITEM 11</b> <b>1 Unit BED, MECHANICAL 2 CRANKS WITH MATTRESS</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 12</b> <b>10 Units PORTABLE UVC LIGHT STERILIZER</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 13</b> <b>3 Unit MEDICINE CABINET WITH GLASS DOOR</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 14</b> <b>1 Unit CAUTERY MACHINE</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 15</b> <b>4 Units INSTRUMENT CABINET</b>	
<b>From</b>	<b>To</b>
Dimensions: 86* 40* 180 cm	Dimensions: Manufacturer's standard
<b>ITEM 16</b> <b>4 UNITS MINOR SURGICAL SET or CUT DOWN SET</b>	
<b>From</b>	<b>To</b>
Haemostatic	Hemostatic
Instrument tray with cover, stainless steel	Instrument tray with cover, stainless steel (Can fit all the instruments inside) (For clarification)
<b>ITEM 17</b> <b>3 Unit EXAMINATION TABLE</b>	
<b>From</b>	<b>To</b>
Material: Stainless	Material: Powder Coated (Hamilton Type) (For clarification)
<b>ITEM 18</b> <b>2 UNITS INSTRUMENT CABINET DOUBLE DOOR</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 19</b> <b>4 UNITS INSTRUMENT TABLE</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 20</b> <b>3 UNITS STERILIZING UNIT</b>	
* No changes as stated in Technical Specifications	
<b>ITEM 21</b> <b>1 UNIT DENTAL INSTRUMENT CABINET</b>	
<b>From</b>	<b>To</b>
Color: 4 color	Color: Manufacturer's standard
<b>ITEM 22</b> <b>1 UNIT SUCTION MACHINE</b>	
* No changes as stated in Technical Specifications	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:



This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:



**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson 



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 48E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipment for the use of Health Facilities in the Mandaluyong City**

This Supplemental/Bid Bulletin No. I is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b> <b>1 UNIT FETAL MONITOR</b>  *No changes as stated in Technical Specifications	
<b>Item 2</b> <b>1 UNIT HEMATOLOGY ANALYZER</b>  *No changes as stated in Technical Specifications	
<b>Item 3</b> <b>1 UNIT MINOR SURGICAL SET</b>	
<b>From</b>	<b>To</b>
Instrument Tray with cover, 18/8, stainless steel (410mm x 254mm x 64mm)	Instrument Tray with cover, stainless
<b>Item 4</b> <b>13 Units DENTAL AEROSOL SUCTION MACHINE</b>  * No changes as stated in Technical Specifications	
<b>ITEM 5</b> <b>7 UNITS EENT DIAGNOSTIC SET</b>	
<b>From</b>	<b>To</b>
Spare light bulb (LED/Xenon)	Spare light bulb (LED/Xenon/Halogen)
<b>ITEM 6</b> <b>11 Units DRESSING SET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 7</b> <b>7 Unit FETAL DOPPLER</b>	
<b>From</b>	<b>To</b>
LCD Display	LCD or LED Display
<b>ITEM 8</b> <b>3 Units INFANT DIGITAL WEIGHING SCALE</b>  If battery operated: Additional Accessory: Rechargeable Batteries: Manufacturer's standard	
<b>ITEM 9</b> <b>1 Unit CERVICAL INSPECTION SET</b>  Additional Accessories:  *Stainless steel tray  * Dimension: Manufacturer's standard (Can fit all the instruments inside the tray)	
<b>ITEM 10</b> <b>1 Unit DENTAL INSTRUMENT CABINET</b>	
<b>From</b>	<b>To</b>
1 Unit DENTAL INSTRUMENT CABINET	1 Unit DENTAL INSTRUMENT CABINET

<b>MATTRESS</b>	<b>MATTRESS (Remove)</b>
Color: 4 color	Color: Manufacturer's standard
<b>ITEM 11</b> <b>3 Units STERILIZING UNIT</b>  * No changes as stated in Technical Specifications	
<b>ITEM 12</b> <b>7 Units MEDICINE CABINET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 13</b> <b>4 Units INSTRUMENT CABINET</b>  * No changes as stated in Technical Specifications	
<b>ITEM 14</b> <b>1 UNIT ELECTRIC DELIVERY TABLE</b>  * No changes as stated in Technical Specifications	
<b>ITEM 15</b> <b>2 Units OPERATING ROOM LIGHT</b>  * No changes as stated in Technical Specifications	
<b>ITEM 16</b> <b>3 UNITS PORTABLE UVC LIGHT STERILIZER</b>  * No changes as stated in Technical Specifications	
<b>ITEM 17</b> <b>3 UNITS INSTRUMENT TABLE</b>	
<b>From</b>	<b>To</b>
Material: Stainless	Material: Stainless steel or equivalent (Remove)
<b>ITEM 18</b> <b>2 UNITS FETAL MONITOR</b>  * No changes as stated in Technical Specifications	
<b>ITEM 19</b> <b>1 UNIT REAGENT REFRIGERATOR</b>  * No changes as stated in Technical Specifications	
<b>ITEM 20</b> <b>1 UNIT ECG MACHINE</b>  With inclusion of: <ul style="list-style-type: none"> <li>Cable and accessories</li> </ul>	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 25<sup>th</sup> day of October 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15<sup>th</sup> day of October, 2021 in MMCHD

Approved by:

  
**ALELI ANNE GRACE P. SUDIACAL, MD, MPH**  
 Director III / BAC Chairperson