



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 037E

PROCUREMENT OF VARIOUS MEDICAL FURNITURE (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1 – 2 UNITS DENTAL INSTRUMENT CABINET	
Material: Stainless steel-at least grade 304	
From	To
Item No. 1 – 2 UNITS DENTAL INSTRUMENT CABINET Size: Manufacturer's Standard	Size: Manufacturer's Standard - removed Dimension: Height: at least 820mm Length/Width: at least 500mm Depth: at least 500mm
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 2 – 6 UNITS INSTRUMENT CABINET	
Stainless steel at least grade 304	
From	To
Item No. 2 – 6 UNITS INSTRUMENT CABINET Three Glass Shelves (adjustable)	Three stainless Steel shelves (adjustable)
Dimensions: Manufacturer's Standard	Dimensions: Height: at least 1800mm Length: at least 800mm Depth: at least 400 mm
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 3 – 5 UNITS INFANT WEIGHING SCALE, DIGITAL	
*No changes as stated in technical specifications.	
Item No. 4 – 7 UNITS MEDICINE CABINET	
Stainless Steel grade at least 304	
Item no. 5 – 2 UNITS ER BED / TRANSFER STRETCHER	
*No changes as stated in technical specifications.	
From	To
Item No. 6 – 8 UNITS DELIVERY TABLE • Dimensions: At least 1900 mm x 550 mm x 800 mm	Dimensions: At least 1800 mm x 550 mm x 800 mm
Item No. 7 – 2 UNITS WEIGHING SCALE WITH HEIGHT MEASURING STICK, HEAVY DUTY, ADULT Weight increments/graduation of 50 grams	Weight increments/graduation of 100 grams



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson 8



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 1	DENTAL INSTRUMENT CABINET	Qty./Unit	2 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 50,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> ● Material: Body -Stainless Steel / Table - Tempered Glass ● Material: Stainless steel-at least grade 304 ● No. of Drawers: At least Five (5) ● Dimension: Height: at least 820mm Length/Width: at least 500mm Depth: at least 500mm ● Accessories: - Silent anti-skid wheel with brakes - Anti-corrosion, fireproof, high hardness, tempered glass top with baffle. - 5 suitable drawer and compartment for different demand - Separating plate 				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to				




METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X</p> <p>Recipient:</p> <p>Ospital ng Paranaque II- 2 units</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 2	INSTRUMENT CABINET	Qty./Unit	6 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 150,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: Stainless steel sheet with two side and back stainless, two glass doors with lock Three stainless Steel shelves (adjustable) Base compartment with two stainless steel doors, one shelf with lock mounted on heavy-duty rubber casters with lock Dimensions: Height: at least 1800mm Length: at least 800mm Depth: at least 400 mm Stainless steel at least grade 304				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no				



3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

Recipient:

C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 3	INFANT WEIGHING SCALE, DIGITAL	Qty./Unit	5 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 75,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: Type: Digital Measuring System: Metric Scale range (Weight): 0-44 lbs / 20 kg in grams Material: ABS Medical Plastic, Detachable Tray with lock Power supply must be 220VAC, 60 Hz or Battery operated Tray Length: At least 50 cm Tray depth: At least 7 cm				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for one (1)				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> a) Service manual in English language b) Operation manual in English language <p>6. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x</p> <p>Recipients:</p> <p>Ospital ng Paranaque I - 2 units</p> <p>Ospital ng Paranaque II - 3 units</p>		
<p>c. <u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 4	MEDICINE CABINET	Qty./Unit	7 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 175,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: Stainless steel sheet with two side stainless steel Stainless Steel grade at least 304 Two glass doors with lock Three glass shelves (adjustable) or three stainless steel shelves (adjustable) Mounted on four swivel caster rubber wheels Dimension: HxLxW; At least 1800mmx800mmx400mm				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X</p> <p>Recipient:</p> <p>Ospital ng Paranaque II - 7 units</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		




Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 5	ER BED / TRANSFER STRETCHER	Qty./Unit	2 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 140,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> ● Safe Working Load: At least 300 kg ● Height Adjustment: Can accommodate from 580 mm to 870 mm ● Trendelenburg/Reverse Trendelenburg: -18° to 18° ● Protection Class: IPX4 or better ● Width Siderails Down: 700mm to 800mm ● Width Siderails Up: 800mm to 850mm ● Total Length: 2100mm to 2200mm ● Mattress Width: 650mm to 700mm ● Mattress Thickness: At least 75mm ● Siderails Height: At least 355mm ● Siderails Length: At least 1190mm ● Features: <ul style="list-style-type: none"> - Made up of steel-reinforced frame. - Fully integrated side rails with gap at the head - With hydraulic control pedals for height adjustment on both sides. - With control pedal for Trendelenburg, Reverse Trendelenburg and Zero leveling on both sides. - With push handle at the head end. - Fold-down side rails. - With four (4) corner roller bumpers. 				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<ul style="list-style-type: none"> - With brake pedals at head and foot end. - Integrated IV pole. ● Standard Accessories: - Oxygen tank holder - IV pole (2 hooks or more) - Transfer Mattress 		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> a) Service manual in English language b) Operation manual in English language <p>6. With "DOH-MMCHD HFEP" (Government</p>		



X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X


Ospital ng Paranaque II - 2 units

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 6	DELIVERY TABLE	Qty./Unit	8 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 480,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> Purpose: Gynecological diagnosis, operation, parturition, clinic Dimensions: At least 1800 mm x 550 mm x 800 mm Backrest lifting: 0-50° Footrest lifting: 0-75° Three sections: Backrest and footrest are adjustable Material: Stainless Steel Table surface: Leather Inner Material: High density sponge With 2 pcs of footplate cover with leather and sponge Standard Accessories <ul style="list-style-type: none"> - 1 pair of footrest - 1 pc mattress - 1 pc dirt bucket - 1 pc sundries shelf 				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to				



2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

- a) Service manual in English language
- b) Operation manual in English language

X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X

Ospital ng Paranaque II – 2 units

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 7	WEIGHING SCALE WITH HEIGHT MEASURING STICK, HEAVY DUTY, ADULT	Qty./Unit	2 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 30,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: Mechanical weighing scale with height rod 220 kgs maximum weighing capacity Non-slip steel platform Adjustable/telescopic height rod with 1 mm graduation Height rod can measure heights from 0.6 meter up to 2 meters high Engraved numbers on eye level scale beam and height rod Large low-rise platform Weight increments/graduation of 100 grams Height rod of aluminum material with plastic head positioner Powder coated body "No sharp edges" construction Built in transport castors				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>6. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x-x</p> <p>Recipient:</p> <p>Ospital ng Paranaque I - 2 units</p>		
<p>c. <u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		