



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB 2022 – 047E**

### PROCUREMENT OF 21 UNITS INDUSTRIAL UV LIGHT

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>
--

*No changes as stated in technical specifications.
--

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18<sup>th</sup> day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of April, 2022 in MMCHD

Approved by:

A handwritten signature in black ink, appearing to read "Aleli Annie Grace P. Sudiacal".

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**

Director III / BAC Chairperson *y*



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>				
<b>Item No. 1</b>	<b>INDUSTRIAL UV LIGHT</b>	Qty./Unit	<b>21 Units</b>	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: <b>1,260,000.00</b>				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
<b>A. Technical Specifications:</b>  <ul style="list-style-type: none"> <li>• Area: At least 60 m2 (unobstructed)</li> <li>• Efficiency: 99.9% effective against harmful germs, viruses, bacteria, etc.</li> <li>• Wavelength: 100-280 nm</li> <li>• Lamp Life: At least 6000 hours</li> <li>• Sterilization Time: 15mins - 24 hrs</li> <li>• Material:               <ul style="list-style-type: none"> <li>- Body: Carbon Steel or better</li> <li>- UV Lamp: Quartz Glass</li> </ul> </li> <li>• Lamp Wattage: 30W*4</li> <li>• Rated Power: At least 120W</li> <li>• Power Supply: 220V, 50/60Hz</li> </ul>				
<b><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></b>  <b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to				





Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		
---	--	--