



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 050E

PROCUREMENT OF 16 UNITS FETAL MONITOR

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:
--

Recipient:

Health Facilities in Paranaque City


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson *af*




Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 1	FETAL MONITOR	Qty./Unit	16 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 2,560,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> Dual pulsed Doppler FHR Ultrasound Frequency: 1 MHz Range: 30-240 bpm, 50-210 bpm With auto-detection of dual fetal movement With array-type thermal printer With automatic printing function Approx. 7-inch. TFT color LCD Display 3-channel display: FHR I, FHR II, UC With Clinical event marking Up to 150 patient data saving Up to 150 patient data saving Dimensions: Manufacturer's Standard Can be connected to a central monitoring system through PC interface Accessories <ul style="list-style-type: none"> - 2 units Ultrasound Doppler probe - 1 unit event marker jack - 3 units probe belt - 1 unit UC probe - 1 unit ultrasound gel - 2 units probe holder 				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- Battery		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2; Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4.Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>7. With "DOH-MMCHD HFEP" (Government</p>		



Property not for sale) sticker in each unit		
x-x Recipient: Health Facilities in Paranaque City		
<u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u> 1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.		