



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 052E

PROCUREMENT OF 39 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Type: Table top	
From	To
Power Supply: 220V, 50/60Hz or Battery operated	To: Power Supply: Dual power operation or autovolt
Safety Standard: EN60601-1	Safety Standard: Manufacturer's standard
Protection Class: BF, Class I	Protection Class: BF, Class II
Compact side configuration	With Handle
Display: LED	Display: LCD or better

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson *yg*



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	PULSE OXIMETER WITH ADULT AND PEDIA PROBE	Qty./Unit	39 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 2,730,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: Type: Table Top Dimensions: Manufacturer's Standard Power Supply: Dual power operation or autovolt Battery Type: Manufacturer's Standard Operation: At least 4 hours Charging Time: 300 min Safety Standard: Manufacturer's standard Protection Class: BF, Class II Conformity: MDD 93/42/EEC, CE0470 or equivalent Enhanced SpO2 Accuracy during motion and low perfusion With Handle 72 hours data memory Built-in rechargeable battery Adult and pediatric application Display: LCD or better			



Audible and visible alarm		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2; Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4.Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>7. With "DOH-MMCHD HFEP" (Government</p>		

